

## Pancreatic amylase CC\* FS\*\*

Diagnostic reagent for quantitative in vitro determination of pancreatic amylase in serum or plasma on DiaSys respons<sup>®</sup>910


### Order Information

Cat. No. 1 0551 99 10 921  
4 twin containers for 120 tests each

### Method

Enzymatic photometric test in which the substrate 4,6-ethylidene-(G7)-p-nitrophenyl-(G1)-α-D-maltoheptaoside (EPS-G7) is cleaved by α-amylases into various fragments. These are further hydrolyzed in a second step by α-glucosidase producing glucose and p-nitrophenol [1,2]. As the salivary isoenzyme is inhibited selectively by a combination of two monoclonal antibodies during the preincubation phase, the increase in absorbance represents the pancreatic amylase activity in the sample [3-5].

### Principle



### Reagents

#### Components and Concentrations

|            |  |         |             |
|------------|--|---------|-------------|
| <b>R1:</b> | Good's buffer  | pH 7.15 | 0.1 mol/L   |
|            | NaCl   |         | 62.5 mmol/L |
|            | MgCl <sub>2</sub>                                      |         | 12.5 mmol/L |
|            | α-Glucosidase  |         | ≥ 2.5 kU/L  |
|            | Monoclonal antibodies against salivary amylase (mouse) |         | ≥ 31 mg/L   |
| <b>R2:</b> | Good's buffer  | pH 7.15 | 0.1 mol/L   |
|            | EPS-G7   |         | 8.5 mmol/L  |

#### Storage Instructions and Reagent Stability

The reagents are stable up to the end of the indicated month of expiry, if stored at 2 – 8°C, protected from light and contamination is avoided. DiaSys respons containers provide protection from light. Do not freeze the reagents!

#### Warnings and Precautions

- The remaining activity of salivary α-amylase is up to 3 %. Very rarely extremely high activities of salivary α-amylase may lead to increased readings of pancreatic α-amylase. However saliva and skin do contain α-amylase, therefore, avoid contact with the reagents.
- The reagents contain sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes.
- Reagent 1 contains animal material. Handle the product as potentially infectious according to universal precautions and good laboratory practice.
- In very rare cases, samples of patients with gammopathy might give falsified results [10].
- Please refer to the safety data sheets and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
- For professional use only!

#### Waste Management

Please refer to local legal requirements.

#### Reagent Preparation

The reagents are ready to use. The bottles are placed directly into the reagent rotor.

#### Specimen

Serum, heparin plasma or EDTA plasma

|                |        |    |           |
|----------------|--------|----|-----------|
| Stability [6]: | 7 days | at | 20 – 25°C |
|                | 7 days | at | 4 – 8°C   |
|                | 1 year | at | -20°C     |

Discard contaminated specimens. Freeze only once.

#### Calibrators and Controls

For calibration, the DiaSys TruCal U calibrator is recommended. This method is traceable to the molar extinction coefficient. For internal quality control DiaSys TruLab N and P controls should be assayed. Each laboratory should establish corrective action in case of deviations in control recovery.

|          | Cat. No.         | Kit size  |
|----------|------------------|-----------|
| TruCal U | 5 9100 99 10 063 | 20 x 3 mL |
|          | 5 9100 99 10 064 | 6 x 3 mL  |
| TruLab N | 5 9000 99 10 062 | 20 x 5 mL |
|          | 5 9000 99 10 061 | 6 x 5 mL  |
| TruLab P | 5 9050 99 10 062 | 20 x 5 mL |
|          | 5 9050 99 10 061 | 6 x 5 mL  |

Reagent information

### Performance Characteristics

|   |                          |
|---|--------------------------|
| Measuring range up to 2000 U/L Pancreatic amylase (in case of higher activities re-measure samples after manual dilution with NaCl solution (9 g/L) or use rerun function). |                          |
| Limit of detection***   | 2 U/L Pancreatic amylase |
| On-board stability  | 6 weeks                  |
| Calibration stability   | 6 weeks                  |

| Interfering substance   | Interferences < 10% | P- amylase [U/L] |
|-------------------------|---------------------|------------------|
| Ascorbate               | up to 30 mg/dL      | 80.8             |
| Hemoglobin              | up to 200 mg/dL     | 54.8             |
|                         | up to 200 mg/dL     | 175              |
| Bilirubin, conjugated   | up to 45 mg/dL      | 54.7             |
|                         | up to 45 mg/dL      | 180              |
| Bilirubin, unconjugated | up to 50 mg/dL      | 55.1             |
|                         | up to 50 mg/dL      | 188              |
| Lipemia (triglycerides) | up to 2000 mg/dL    | 51.7             |
|                         | up to 1200 mg/dL    | 241              |

For further information on interfering substances refer to Young DS [7].

| Precision                    |          |          |          |
|------------------------------|----------|----------|----------|
| Within run (n=20)            | Sample 1 | Sample 2 | Sample 3 |
| Mean [U/L]                   | 31.9     | 143      | 295      |
| Coefficient of variation [%] | 1.80     | 2.43     | 1.95     |
| Between run (n=20)           | Sample 1 | Sample 2 | Sample 3 |
| Mean [U/L]                   | 25.2     | 102      | 133      |
| Coefficient of variation [%] | 3.44     | 2.31     | 1.99     |

| Method comparison (n=137)  |  |
|----------------------------|--|
| Test x                     | DiaSys Pancreatic amylase CC FS (Hitachi 917)              |
| Test y                     | DiaSys Pancreatic amylase CC FS (respons <sup>®</sup> 910) |
| Slope                      | 0.959  |
| Intercept                  | 0.349 U/L  |
| Coefficient of correlation | 0.99998  |

\*\*\* according to NCCLS document EP17-A, vol. 24, no. 34

#### Conversion factor

Pancreatic amylase [U/L] x 0.0167 = Pancreatic amylase [μkat/L]

#### Reference Range [8]


|              |               |               |
|--------------|---------------|---------------|
| Serum/plasma | Women         | Men           |
|              | < 53 U/L      | < 53 U/L      |
|              | < 0.88 μkat/L | < 0.88 μkat/L |

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

#### Literature

- Lorentz K. α-Amylase. In: Thomas L, editor. Clinical laboratory diagnostics. 1st ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 192-202.
- Moss DW, Henderson AR. Digestive enzymes of pancreatic origin. In: Burtis CA, Ashwood ER, editors. Tietz Textbook of Clinical Chemistry. 3rd ed. Philadelphia: W.B Saunders Company; 1999. p. 689-98.
- Gerber M, Naujoks K, Lenz H, Wulff K. A monoclonal antibody that specifically inhibits human salivary alpha-amylase. Clin Chem 1987; 33: 1158-62.
- Kruse-Jarres JD, Kaiser C, Hafkenscheid JC, Hohenwallner W, Stein W., Bohner J et al. Evaluation of a new alpha-amylase assay using 4,6-ethylidene-(G7)-1-4-nitrophenyl-(G1)-alpha-D-maltoheptaoside as substrate. J Clin Chem Biochem 1989; 27: 103-13.
- Tietz NW, Burlina A, Gerhardt W, Junge W, Maffertheimer P, Mural T et al. Multicenter evaluation of a specific pancreatic isoamylase assay based on a double monoclonal-antibody technique. Clin Chem 1988; 34: 2096-102.
- Guder WG, Zawta B et al. The Quality of Diagnostic Samples. 1<sup>st</sup> ed. Darmstadt: GIT Verlag; 2001. p. 16-17.
- Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th. ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press, 2000.
- Junge W, Wortmann W, Wilke B, Waldenstroem J et al. Development and evaluation of assays for determination of total and pancreatic amylase at 37 °C according to the principle recommended by the IFCC. Clin Biochem 2001; 34: 607-15.
- Junge W, Troge B, Klein G, Poppe W, Gerber M. Evaluation of a new assay for pancreatic amylase: Performance characteristics and estimation of reference interval. Clin Biochem 1989; 22: 109-14.
- Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: Mechanisms, detection and prevention. Clin Chem Lab Med 2007; 45(9): 1240-1243.

#### Manufacturer

 DiaSys Diagnostic Systems GmbH  
Alte Strasse 9 65558 Holzheim Germany

\*complete color \*\* fluid stable

## Pancreatic amylase CC FS

### Application for serum and plasma samples

This application was set up and evaluated by DiaSys. It is based on the standard equipment at that time and does not apply to any equipment modifications undertaken by unqualified personnel.

| Identification                      |      |
|-------------------------------------|------|
| This method is usable for analysis: | Yes  |
| Twin reaction:                      | No   |
| Name:                               | PAMY |
| Shortcut:                           |      |
| Reagent barcode reference:          | 016  |
| Host reference:                     | 016  |

| Technic                               |                |
|---------------------------------------|----------------|
| Type:                                 | Linear kinetic |
| First reagent:[ $\mu$ L]              | 160            |
| Blank reagent                         | Yes            |
| Sensitive to light                    |                |
| Second reagent:[ $\mu$ L]             | 40             |
| Blank reagent                         | No             |
| Sensitive to light                    |                |
| Main wavelength:[nm]                  | 405            |
| Secondary wavelength:[nm]             | 700            |
| Polychromatic factor:                 | 1.0000         |
| 1 st reading time [min:sec]           | 07:48          |
| Last reading time [min:sec]           | 10:00          |
| Reaction way:                         | Increasing     |
| Linear Kinetics                       |                |
| Substrate depletion: Absorbance limit | 1.2000         |
| Linearity: Maximum deviation [%]      | 100.0000       |
| Fixed Time Kinetics                   |                |
| Substrate depletion: Absorbance limit |                |
| Endpoint                              |                |
| Stability: Largest remaining slope    |                |
| Prozone Limit [%]                     |                |

| Reagents |  |
|----------|--|
| Decimals |  |
| Units    |  |

| Sample                               |                  |
|--------------------------------------|------------------|
| Diluent                              | DIL A (NaCl)     |
| Hemolysis:                           |                  |
| Agent [ $\mu$ L]                     | 0 (no hemolysis) |
| Cleaner                              |                  |
| Sample [ $\mu$ L]                    | 0                |
| Technical limits                     |                  |
| Concentration technical limits-Lower | 2.0000           |
| Concentration technical limits-Upper | 2000.0000        |
| SERUM                                |                  |
| Normal volume [ $\mu$ L]             | 4.0              |
| Normal dilution (factor)             | 1                |
| Below normal volume [ $\mu$ L]       |                  |
| Below normal dilution (factor)       |                  |
| Above normal volume [ $\mu$ L]       | 4.0              |
| Above normal dilution (factor)       | 6                |
| URINE                                |                  |
| Normal volume [ $\mu$ L]             | 4.0              |
| Normal dilution (factor)             | 1                |
| Below normal volume [ $\mu$ L]       |                  |
| Below normal dilution (factor)       |                  |
| Above normal volume [ $\mu$ L]       | 4.0              |
| Above normal dilution (factor)       | 6                |
| PLASMA                               |                  |
| Normal volume [ $\mu$ L]             | 4.0              |
| Normal dilution (factor)             | 1                |
| Below normal volume [ $\mu$ L]       |                  |
| Below normal dilution (factor)       |                  |
| Above normal volume [ $\mu$ L]       | 4.0              |
| Above normal dilution (factor)       | 6                |
| CSF                                  |                  |
| Normal volume [ $\mu$ L]             | 4.0              |
| Normal dilution (factor)             | 1                |
| Below normal volume [ $\mu$ L]       |                  |
| Below normal dilution (factor)       |                  |
| Above normal volume [ $\mu$ L]       | 4.0              |
| Above normal dilution (factor)       | 6                |
| Whole blood                          |                  |
| Normal volume [ $\mu$ L]             | 4.0              |
| Normal dilution (factor)             | 1                |
| Below normal volume [ $\mu$ L]       |                  |
| Below normal dilution (factor)       |                  |
| Above normal volume [ $\mu$ L]       | 4.0              |
| Above normal dilution (factor)       | 6                |

| Results                   |        |
|---------------------------|--------|
| Decimals                  | 1      |
| Units                     | U/L    |
| Correlation factor-Offset | 0.0000 |
| Correlation factor-Slope  | 1.0000 |

| Range       |           |
|-------------|-----------|
| Gender      | All       |
| Age         |           |
| SERUM       | >= <=53.0 |
| URINE       |           |
| PLASMA      | >= <=53.0 |
| CSF         |           |
| Whole blood |           |
| Gender      |           |
| Age         |           |
| SERUM       |           |
| URINE       |           |
| PLASMA      |           |
| CSF         |           |
| Whole blood |           |

| Contaminants                              |  |
|---|--|
| Please refer to r910 Carryover Pair Table |  |

| Calibrators details |                |  |
|---------------------|----------------|--|
| Calibrator list     | Concentration  |  |
| Cal. 1/Blank        | 0              |  |
| Cal. 2              | *              |  |
| Cal. 3              |                |  |
| Cal. 4              |                |  |
| Cal. 5              |                |  |
| Cal. 6              |                |  |
|                     | Max delta abs. |  |
| Cal. 1              | 0.002          |  |
| Cal. 2              | 0.005          |  |
| Cal. 3              |                |  |
| Cal. 4              |                |  |
| Cal. 5              |                |  |
| Cal. 6              |                |  |
| Drift limit [%]     | 0.80           |  |

| Calculations |   |
|--------------|---|
| Model        | X |
| Degree       | 1 |

\* Enter calibrator value