

# Immunoglobulin A FS\*

Diagnostic reagent for quantitative in vitro determination of immunoglobulin A (IgA) in serum or plasma on photometric systems

## Order Information

Cat. No.	Kit size
1 7202 99 10 930	R1 4 x 20 mL + R2 2 x 8 mL
1 7202 99 10 935	R1 2 x 20 mL + R2 1 x 8 mL
1 7202 99 90 309	R1 4 x 20 mL + R2 2 x 8 mL
5 9200 99 10 037	3 x 1 mL TruCal Protein high
5 9200 99 10 039	5 x 1 mL TruCal Protein:

Calibrator set with 5 different levels

## Summary [1-3]

The human immunoglobulin classes (IgG, IgA, IgM, IgE and IgD) are a group of functionally and structurally closely related glycoproteins. Human IgA has a molecular weight of about 160 000 dalton and consists of two identical heavy chains and two identical light chains which are bound together by disulfide bonds in a characteristic Y-shaped form. Serum IgA is produced by plasma cells (B-cells) and represents about 15% of all soluble immunoglobulin classes. About 90% of the serum IgA is monomeric the rest is dimeric and polymeric. Most of IgA is not present in serum but on the surface of mucous membranes. In the mucosal tissues of the lung and the gastrointestinal tract IgA is released by plasma cells in a dimeric form. The two Y-shaped pieces are bound together not only by a joining chain but also by a special peptide called secretory component. This IgA type is called secretory-IgA. It is normally not present in human serum but in other body fluids like sweat, tears, gastrointestinal and bronchial secretions. The main function of serum-IgA is to bind to antigens and trigger further catabolism of the antigen.

Decreased serum-IgA concentrations occur in primary as well as in secondary immunodeficiency syndromes. A high increase of one immunoglobulin class due to multiple myeloma may result in a decrease in other immunoglobulin classes like IgA. Increased loss of IgA due to severe enteritis may result in a decreased concentration. Increased IgA concentrations can be observed in severe infections and autoimmune diseases. Especially inflammatory processes of the liver may result in increased serum IgA levels. Like for other Ig-classes many forms of myeloma produce high amounts of monoclonal or polyclonal IgA. Quantitative serum-IgA determination is necessary for differential diagnostics of these diseases.

All methods for IgA quantitation are calibrated for polyclonal serum-IgA. The quantitation of monoclonal IgA is not standardized and values may differ for different reagents and methods. Values should only be used for follow up studies. Monoclonal immunoglobulinemia requires detailed differential diagnostic investigation in addition to the quantitative determination.

## Method

Immunturbidimetric test

## Principle

Determination of IgA concentration by photometric measurement of antigen-antibody-reaction of antibodies to human IgA with IgA present in the sample.

## Reagents

### Components and Concentrations

<b>R1:</b>	TRIS	pH 7.5	100 mmol/L
	NaCl		150 mmol/L
<b>R2:</b>	TRIS	pH 8.0	100 mmol/L
	NaCl		300 mmol/L
	Anti-human IgA antibody (goat)		< 1%

### Storage Instructions and Reagent Stability

The reagents are stable up to the end of the indicated month of expiry, if stored at 2 – 8 °C and contamination is avoided. Do not freeze the reagents and store them protected from light!

### Warnings and Precautions

1. The reagents contain sodium azide (0.95 g/L) as preservative. Do not swallow! Avoid contact with skin and mucous membranes!
2. Reagent 2: contains animal material. Handle the product as potentially infectious according to universal precautions and good clinical laboratory practices.
3. In very rare cases, samples of patients with gammopathy might give falsified results [8].
4. Please refer to the safety data sheets and take the necessary precautions for the use of laboratory reagents. For diagnostic purposes, the results should always be assessed with the patient's medical history, clinical examinations and other findings.
5. For professional use only!

### Waste Management

Please refer to local legal requirements.

### Reagent Preparation

The reagents are ready to use.

### Materials required but not provided

NaCl solution 9 g/L  
General laboratory equipment

### Specimen

Serum, heparin plasma or EDTA plasma		
Stability [4]:	3 months at	20 – 25°C
	3 months at	4 – 8°C
	6 months at	-20°C

Only freeze once!

Discard contaminated specimens.

### Assay Procedure for Analyzers

**Application sheets for automated systems are available on request.**

Wavelength	570 nm
Optical path	1 cm
Temperature	37°C
Measurement	Against reagent blank

	Blank	Sample or calibrator
Sample or calibrator	-	2 µL
Dist. water	2 µL	-
Reagent 1	250 µL	250 µL
Mix, incubate for 3 – 5 min., read absorbance (A1), then add:		
Reagent 2	50 µL	50 µL
Mix, incubate for 3 min., read absorbance (A2).		

$$\Delta A = (A2 - A1) \text{ sample or calibrator}$$

## Calculation

The concentration of IgA in unknown samples is derived from a calibration curve using an appropriate mathematical model such as logit/log. The calibration curve is obtained with 5 calibrators at different levels and NaCl solution (9 g/L) for determination of the zero value.

Stability of calibration: 4 weeks

## Conversion factor

Immunoglobulin A [mg/dL] x 0.0625 = Immunoglobulin A [ $\mu$ mol/L]

## Calibrators and Controls

For the calibration of automated photometric systems, DiaSys' TruCal Protein calibrator set or TruCal Protein high calibrator are recommended. The assigned values of the calibrators have been made traceable to the reference material ERM<sup>®</sup>-DA470k/IFCC. For internal quality control a DiaSys TruLab Protein control should be assayed. Each laboratory should establish corrective action in case of deviations in control recovery.

	Cat. No.	Kit size
TruLab Protein Level 1	5 9500 99 10 046	3 x 1 mL
TruLab Protein Level 2	5 9510 99 10 046	3 x 1 mL

## Performance Characteristics

### Measuring Range

The test has been developed to determine concentrations of IgA within a measuring range from 30 - 900 mg/dL, at least up to the concentration of the highest calibrator. When values exceed the upper range, samples should be diluted 1 + 1 with NaCl solution (9 g/L) and the result multiplied by 2.

When results are below the lower range, repeat measurement with double sample volume. If results are still below the range, check for prozone effect by diluting the sample.

### Prozone Limit

No prozone effect was observed up to an IgA value of 5000 mg/dL.

### Specificity/Interferences

Due to its antibodies, DiaSys Immunoglobulin A FS is a specific immunoassay for human IgA. No interference was observed by conjugated and unconjugated bilirubin up to 60 mg/dL, hemoglobin up to 1000 mg/dL, lipemia up to 2000 mg/dL triglycerides and RF up to 1700 IU/mL.

No cross reaction with IgG or IgM was observed under test conditions. For further information on interfering substances refer to Young DS [5].

### Sensitivity/Limit of Detection

The lower limit of detection (the minimum concentration which can be measured and distinguished from zero) is 8 mg/dL.

### Imprecision

According to protocol EP-5 of the NCCLS (National Committee of Clinical Laboratory Standards)

Within-run precision n = 40	Mean [mg/dL]	SD [mg/dL]	CV [%]
Sample 1	296	7.68	2.60
Sample 2	407	10.9	2.68
Sample 3	499	10.7	2.14

Between day precision n = 40	Mean [mg/dL]	SD [mg/dL]	CV [%]
Sample 1	296	3.90	1.32
Sample 2	407	3.40	0.83
Sample 3	499	8.43	1.69

## Method Comparison

A comparison of DiaSys Immunoglobulin A FS (y) to an immunoturbidimetric test (x) using 81 samples gave following results:

$$y = 0.86 x + 19.1 \text{ mg/dL}; r = 0.983.$$

A comparison of DiaSys Immunoglobulin A FS (y) to a nephelometric test (x) using 79 samples gave following results:

$$y = 1.07 x + 0.18 \text{ mg/dL}; r = 0.996.$$

## Reference values

Adults [6]	70 – 400 mg/dL	4.38 – 25.0 $\mu$ mol/L	
Children [7]	< 1 month	7 – 94 mg/dL	0.44 – 5.88 $\mu$ mol/L
	1 – 12 month/s	10 – 131 mg/dL	0.63 – 8.19 $\mu$ mol/L
	1 – 3 year/s	19 – 220 mg/dL	1.19 – 13.8 $\mu$ mol/L
	4 – 5 years	48 – 345 mg/dL	3.00 – 21.6 $\mu$ mol/L
	6 – 7 years	41 – 297 mg/dL	2.56 – 18.6 $\mu$ mol/L
	8 – 10 years	51 – 297 mg/dL	3.19 – 18.6 $\mu$ mol/L
	11 – 13 years	44 – 395 mg/dL	2.75 – 24.7 $\mu$ mol/L

Each laboratory should check if the reference ranges are transferable to its own patient population and determine own reference ranges if necessary.

## Literature

1. Thomas L. Clinical Laboratory Diagnostics. 1<sup>st</sup> ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 667-78.
2. Johnson AM, Rohlf EM, Silverman LM. Proteins. In: Burtis CA, Ashwood ER. editors. Tietz textbook of clinical chemistry. 3<sup>rd</sup> ed. Philadelphia: W. B. Saunders Company; 1999. p. 507-12.
3. Bartl R, Hoechtlen-Vollmar W, Thomas L. Monoclonal immunoglobulins. In: Thomas L. Clinical Laboratory Diagnostics. 1<sup>st</sup> ed. Frankfurt: TH-Books Verlagsgesellschaft; 1998. p. 742-58.
4. Guder WG, Narayanan S et al. List of Analytes; Preanalytical Variables. 1<sup>st</sup> ed. Darmstadt: Git Verlag, 1996: 16-7.
5. Young DS. Effects of Drugs on Clinical Laboratory Tests. 5th ed. Volume 1 and 2. Washington, DC: The American Association for Clinical Chemistry Press 2000.
6. Dati F, Schumann G, Thomas L, Aguzzi F, Baudner S, Bienvenu J et al. Consensus of a group of professional societies and diagnostic companies on guidelines for interim reference ranges for 14 proteins in serum based on the standardization against the IFCC/BCR/CAP reference material (CRM 470). Eur J Clin Chem Clin Biochem 1996;34:517-20.
7. Heil R, Koberstein R, Zawta B. Referenzbereiche für Kinder und Erwachsene. Roche Diagnostics 2004. p. 44-45.
8. Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: Mechanism, detection and prevention. Clin Chem Lab Med 2007; 45(9): 1240-1243.

## Manufacturer



DiaSys Diagnostic Systems GmbH  
Alte Strasse 9 65558 Holzheim Germany