TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric	1 0242	9.82	7.07 - 12.6	mg/L
Microalbumin			0.982	0.707 - 1.26	mg/dL
AMY	EPS G-7	1 0501	40.7	32.6 - 48.9	U/L
Amylase			0.679	0.543 - 0.815	µkat/L
BUN	Urease UV	1 3101	396	293 - 499	mg/dL
Urea N			141	104 - 178	mmol/L
CAAS	Arsenazo	1 1130	1.91	1.58 - 2.23	mmol/L
Calcium			7.65	6.35 - 8.95	mg/dL
CA	Phosphonazo	1 1181	2.08	1.73 - 2.44	mmol/L
Calcium			8.35	6.93 - 9.77	mg/dL
CREA	Jaffé, without compensation	1 1711	66.0	50.2 - 81.9	mg/dL
Creatinine			5.84	4.44 - 7.24	mmol/L
CREAP	Enzymatic PAP	1 1759	70.8	53.8 - 87.8	mg/dL
Creatinine			6.26	4.76 - 7.76	mmol/L
D	Refractrometer		1.011	1.000 - 1.020	
Specific Gravity					
GLUHK	HK/G6P-DH	1 2511	24.2	18.9 - 29.6	mg/dL
Glucose			1.35	1.05 - 1.64	mmol/L
hCG				Negative	
Human Chorionic Gonadotropin					
MG	Xylidyl blue	1 4610	2.29	1.69 - 2.88	mmol/L
Magnesium			5.56	4.11 - 7.00	mg/dL
OSMO	Freezing point		392	314 - 471	mOsm/kg
Osmolality					
pH	pH Meter		6.21	5.50 - 6.90	
PO4	Molybdate UV, substrate start	1 5211	26.8	20.9 - 32.7	mg/dL
Phosphorus			8.65	6.75 - 10.6	mmol/L
TP	Pyrogallol red	1 0210	8.01	6.09 - 9.93	mg/dL
Total Protein			80.1	60.9 - 99.3	mg/L
UREA	Urease UV	1 3101	848	628 - 1068	mg/dL
Urea			141	104 - 178	mmol/L
UA TBH	TBHBA	1 3021	6.21	4.60 - 7.83	mg/dL
Uric Acid			370	273 - 466	µmol/L
UA	TOOS	1 3001	7.08	5.24 - 8.91	mg/dL
Uric Acid			421	311 - 530	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

respons®910

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric	1 0242	9.82	7.07 - 12.6	mg/L
Microalbumin			0.982	0.707 - 1.26	mg/dL
AMY	EPS G-7	1 0501	40.7	32.6 - 48.9	U/L
Amylase			0.679	0.543 - 0.815	µkat/L
BUN	Urease UV	1 3101	396	293 - 499	mg/dL
Urea N			141	104 - 178	mmol/L
CA	Phosphonazo	1 1181	2.08	1.73 - 2.44	mmol/L
Calcium			8.35	6.93 - 9.77	mg/dL
CREA	Jaffé , without compensation	1 1711	66.0	50.2 - 81.9	mg/dL
Creatinine			5.84	4.44 - 7.24	mmol/L
CREAP	Enzymatic PAP	1 1759	70.8	53.8 - 87.8	mg/dL
Creatinine			6.26	4.76 - 7.76	mmol/L
GLUHK	HK/G6P-DH	1 2511	24.2	18.9 - 29.6	mg/dL
Glucose			1.35	1.05 - 1.65	mmol/L
MG	Xylidyl blue	1 4610	2.29	1.69 - 2.88	mmol/L
Magnesium			5.56	4.11 - 7.00	mg/dL
PO4	Molybdate UV	1 5211	26.8	20.9 - 32.7	mg/dL
Phosphorus			8.65	6.75 - 10.6	mmol/L
TP	Pyrogallol red	1 0210	8.01	6.09 - 9.93	mg/dL
Total Protein			80.1	60.9 - 99.3	mg/L
UREA	Urease UV	1 3101	848	628 - 1068	mg/dL
Urea			141	104 - 178	mmol/L
UA	TOOS	1 3001	7.08	5.24 - 8.91	mg/dL
Uric Acid			421	311 - 530	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

respons®920

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	1 SD	Unit
ALB	Immunoturbidimetric	1 0242	9.82	7.07 - 12.6	0.916	mg/L
Microalbumin			0.982	0.707 - 1.26	0.092	mg/dL
AMY	EPS G-7	1 0501	40.7	32.6 - 48.9	2.72	U/L
Amylase			0.679	0.543 - 0.815	0.045	µkat/L
BUN	Urease UV	1 3101	396	293 - 499	34.3	mg/dL
Urea N			141	104 - 178	12.2	mmol/L
CA	Phosphonazo	1 1181	2.08	1.73 - 2.44	0.118	mmol/L
Calcium			8.35	6.93 - 9.77	0.473	mg/dL
CL	Direct ISE	960405	51.8	44.6 - 59.1	2.42	mmol/L
Chloride			184	158 - 210	8.58	mg/dL
CREA	Jaffé , without compensation	1 1711	66.0	50.2 - 81.9	5.28	mg/dL
Creatinine			5.84	4.44 - 7.24	0.467	mmol/L
CREAP	Enzymatic PAP	1 1759	70.8	53.8 - 87.8	5.66	mg/dL
Creatinine			6.26	4.76 - 7.76	0.501	µmol/L
GLUHK	HK/G6P-DH	1 2511	24.2	18.9 - 29.6	1.78	mg/dL
Glucose			1.35	1.05 - 1.65	0.099	mmol/L
MG	Xylidyl blue	1 4610	2.29	1.69 - 2.88	0.198	mmol/L
Magnesium			5.56	4.11 - 7.00	0.482	mg/dL
PO4	Molybdate UV	1 5211	26.8	20.9 - 32.7	1.96	mg/dL
Phosphorus			8.65	6.75 - 10.6	0.634	mmol/L
K	Direct ISE	960403	30.8	25.5 - 36.0	1.74	mmol/L
Potassium			120	99.8 - 141	6.81	mg/dL
Na	Direct ISE	960402	73.4	63.8 - 82.9	3.18	mmol/L
Sodium			169	147 - 191	7.31	mg/dL
UREA	Urease UV	1 3101	848	628 - 1068	73.5	mg/dL
Urea			141	104 - 178	12.2	mmol/L
UA	TOOS	1 3001	7.08	5.24 - 8.91	0.613	mg/dL
Uric Acid			421	311 - 530	36.5	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

respons®940

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	1 SD	Unit
ALB	Immunoturbidimetric	1 0242	9.82	7.07 - 12.6	0.916	mg/L
Microalbumin			0.982	0.707 - 1.26	0.092	mg/dL
AMY	EPS G-7	1 0501	40.7	32.6 - 48.9	2.72	U/L
Amylase			0.679	0.543 - 0.815	0.045	µkat/L
BUN	Urease UV	1 3101	396	293 - 499	34.3	mg/dL
Urea N			141	104 - 178	12.2	mmol/L
CA	Phosphonazo	1 1181	2.08	1.73 - 2.44	0.118	mmol/L
Calcium			8.35	6.93 - 9.77	0.473	mg/dL
CL	Direct ISE	960405	51.8	44.6 - 59.1	2.42	mmol/L
Chloride			184	158 - 210	8.58	mg/dL
CREA	Jaffé , without compensation	1 1711	66.0	50.2 - 81.9	5.28	mg/dL
Creatinine			5.84	4.44 - 7.24	0.467	mmol/L
CREAP	Enzymatic PAP	1 1759	70.8	53.8 - 87.8	5.66	mg/dL
Creatinine			6.26	4.76 - 7.76	0.501	mmol/L
GLUHK	HK/G6P-DH	1 2511	24.2	18.9 - 29.6	1.78	mg/dL
Glucose			1.35	1.05 - 1.65	0.099	mmol/L
MG	Xylidyl blue	1 4610	2.29	1.69 - 2.88	0.198	mmol/L
Magnesium			5.56	4.11 - 7.00	0.482	mg/dL
PO4	Molybdate UV	1 5211	26.8	20.9 - 32.7	1.96	mg/dL
Phosphorus			8.65	6.75 - 10.6	0.634	mmol/L
K	Direct ISE	960403	30.8	25.5 - 36.0	1.74	mmol/L
Potassium			120	99.8 - 141	6.81	mg/dL
Na	Direct ISE	960402	73.4	63.8 - 82.9	3.18	mmol/L
Sodium			169	147 - 191	7.31	mg/dL
TP	Pyrogallol red	1 0210	8.01	6.09 - 9.93	0.641	mg/dL
Total Protein			80.1	60.9 - 99.3	6.41	mg/L
UREA	Urease UV	1 3101	848	628 - 1068	73.5	mg/dL
Urea			141	104 - 178	12.2	mmol/L
UA	TOOS	1 3001	7.08	5.24 - 8.91	0.613	mg/dL
Uric Acid			421	311 - 530	36.5	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

BioMajesty®

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	1 SD	Unit
ALB	Immunoturbidimetric	1 0242	9.82	7.07 - 12.6	0.916	mg/L
Microalbumin			0.982	0.707 - 1.26	0.092	mg/dL
AMY	EPS G-7	1 0501	40.7	32.6 - 48.9	2.72	U/L
Amylase			0.679	0.543 - 0.815	0.045	µkat/L
BUN	Urease UV	1 3101	396	293 - 499	34.3	mg/dL
Urea N			141	104 - 178	12.2	mmol/L
CA	Phosphonazo	1 1181	2.08	1.73 - 2.44	0.118	mmol/L
Calcium			8.35	6.93 - 9.77	0.473	mg/dL
CL	Indirect ISE	980132	60.4	51.9 - 68.8	2.82	mmol/L
Chloride			214	184 - 244	9.99	mg/dL
CREA	Jaffé, without compensation	1 1711	66.0	50.2 - 81.9	5.28	mg/dL
Creatinine			5.84	4.44 - 7.24	0.467	mmol/L
CREAP	Enzymatic PAP	1 1759	70.8	53.8 - 87.8	5.66	mg/dL
Creatinine			6.26	4.76 - 7.76	0.501	mmol/L
GLUHK	HK/G6P-DH	1 2511	24.2	18.9 - 29.6	1.78	mg/dL
Glucose			1.35	1.05 - 1.65	0.099	mmol/L
MG	Xylidyl blue	1 4610	2.29	1.69 - 2.88	0.198	mmol/L
Magnesium			5.56	4.11 - 7.00	0.482	mg/dL
PO4	Molybdate UV	1 5211	26.8	20.9 - 32.7	1.96	mg/dL
Phosphorus			8.65	6.75 - 10.6	0.634	mmol/L
K	Indirect ISE	980131	31.7	26.4 - 37.1	1.80	mmol/L
Potassium			124	103 - 145	7.03	mg/dL
Na	Indirect ISE	980130	81.9	71.3 - 92.5	3.55	mmol/L
Sodium			188	164 - 213	8.16	mg/dL
TP	Pyrogallol red	1 0210	8.01	6.09 - 9.93	0.641	mg/dL
Total Protein			80.1	60.9 - 99.3	6.41	mg/L
UREA	Urease UV	1 3101	848	628 - 1068	73.5	mg/dL
Urea			141	104 - 178	12.2	mmol/L
UA	TOOS	1 3001	7.08	5.24 - 8.91	0.613	mg/dL
Uric Acid			421	311 - 530	36.5	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

ABBOTT AEROSET/ARCHITECT

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Turbidimetric / Immunoturbidimetric	2K98	20.0	14.4 - 25.6	mg/L
Microalbumin			2.00	1.44 - 2.56	mg/dL
AMY	CNPG3 37° C	7D58	*	-	U/L
Amylase				-	µkat/L
	CNPG3 37° C, IFCC	7D58	*	-	U/L
				-	µkat/L
CAAS	Arsenazo III	3L79	2.01	1.67 - 2.35	mmol/L
Calcium			8.06	6.69 - 9.44	mg/dL
		4S91	1.88	1.56 - 2.20	mmol/L
			7.51	6.23 - 8.79	mg/dL
CL	ISE Indirect	2P32	52.0	44.7 - 59.3	mmol/L
Chloride			184	159 - 210	mg/dL
CORT	CMIA	8D15	*	-	µg/dL
Cortisol				-	nmol/L
CREA	Alkaline Picrate	3L81	*	-	mg/dL
Creatinine				-	mmol/L
	Enzymatic	8L24	70.5	53.6 - 87.4	mg/dL
			6.23	4.74 - 7.73	mmol/L
GLUC	Hexokinase/G-6-PDH	3L82	24.1	18.8 - 29.4	mg/dL
Glucose			1.34	1.04 - 1.63	mmol/L
MG	Arsenazo III	7D70	*	-	mmol/L
Magnesium				-	mg/dL
PO4	Phosphomolybdate	7D71	*	-	mg/dL
Phosphorus				-	mmol/L
	Phosphomolybdate	4T07	24.0	18.7 - 29.3	mg/dL
			7.74	6.04 - 9.45	mmol/L
K	ISE Indirect	2P32	31.9	26.5 - 37.3	mmol/L
Potassium			125	104 - 146	mg/dL
Na	ISE Indirect	2P32	81.3	70.7 - 91.9	mmol/L
Sodium			187	163 - 211	mg/dL
TP	Benzethonium Chloride	7D79	7.55	5.74 - 9.36	mg/dL
Total Protein			75.5	57.4 - 93.6	mg/L
UREA	Urease/GLDH	7D75	*	-	mg/dL
Urea				-	mmol/L
	Urease/GLDH	4T12	873	646 - 1100	mg/dL
			146	108 - 184	mmol/L
UA	Uricase (UAU)	3P39	*	-	mg/dL
Uric Acid				-	µmol/L
	Uricase (Uric-U)	4T13	6.00	4.44 - 7.56	mg/dL
			357	264 - 450	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

ABBOTT ALINITY

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric	08P04	18.4	13.2 - 23.6	mg/L
Microalbumin			1.84	1.32 - 2.36	mg/dL
AMY	Enzymatic, CNPG3 37° C	07P58	44.7	35.8 - 53.6	U/L
Amylase			0.745	0.596 - 0.894	µkat/L
	Enzymatic, CNPG3 37° C, IFCC	07P58	*	-	U/L
				-	µkat/L
BUN	Urease/GLDH	08P16	352	260 - 444	mg/dL
Urea N			126	93.0 - 158	mmol/L
CA	Arsenazo III	07P57	1.99	1.65 - 2.33	mmol/L
Calcium			7.94	6.59 - 9.29	mg/dL
CL	ISE Indirect	07P53	52.3	45.0 - 59.6	mmol/L
Chloride			185	159 - 211	mg/dL
CORT	CMIA	08P33	1.10	0.880 - 1.32	µg/dL
Cortisol			30.4	24.3 - 36.5	nmol/L
CREA	Alkaline Picrate	07P99	72.4	55.0 - 89.8	mg/dL
Creatinine			6.40	4.86 - 7.94	mmol/L
	Enzymatic	08P01	71.1	54.0 - 88.2	mg/dL
			6.28	4.77 - 7.79	mmol/L
GLUC	Hexokinase/G-6-PDH	07P55	25.2	19.7 - 30.7	mg/dL
Glucose			1.40	1.09 - 1.71	mmol/L
MG	Enzymatic	08P19	2.41	1.78 - 3.04	mmol/L
Magnesium			5.87	4.34 - 7.40	mg/dL
PO4	Phosphomolybdate	04U03	22.8	17.8 - 27.8	mg/dL
Phosphorus			7.36	5.74 - 8.98	mmol/L
K	ISE Indirect	07P53	32.1	26.6 - 37.6	mmol/L
Potassium			126	104 - 147	mg/dL
Na	ISE Indirect	07P53	84.0	73.1 - 94.9	mmol/L
Sodium			193	168 - 218	mg/dL
TP	Benzethonium Chloride	07P59	7.00	5.32 - 8.68	mg/dL
Total Protein			70.0	53.2 - 86.8	mg/L
UREA	Urease/GLDH	08P16	755	559 - 951	mg/dL
Urea			126	93.0 - 158	mmol/L
UA	Uricase	08P56	6.49	4.80 - 8.18	mg/dL
Uric Acid			386	286 - 486	µmol/L
		04U09	6.02	4.45 - 7.59	mg/dL
			358	265 - 451	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

CE IVD

BECKMAN COULTER ACCESS

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CORT	Chemiluminescence	Cort	2.34	1.87 - 2.81	µg/dL
Cortisol			64.7	51.8 - 77.6	nmol/L

BECKMAN COULTER DXI

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CORT	Chemiluminescence	Cort	*	-	µg/dL
Cortisol				-	nmol/L

ORTHO-VITROS IMMUNOASSAY SYSTEMS

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CORT	Chemiluminescence		*	-	µg/dL
Cortisol				-	nmol/L

SIEMENS CENTAUR

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CORT	Chemiluminescence		*	-	µg/dL
Cortisol				-	nmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

BECKMAN COULTER AU

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB Microalbumin	Turbidimetric	OSR6167	*	-	mg/L
				-	mg/dL
	Turbidimetric (UALB)		15.1	10.9 - 19.3	mg/L
			1.51	1.09 - 1.93	mg/dL
AMY Amylase	CNPG3	OSR6x06	35.1	28.1 - 42.1	U/L
			0.585	0.468 - 0.702	µkat/L
	G7PNP/IFCC	OSR6182	*	-	U/L
				-	µkat/L
BUN Urea N	Urease/GLDH	OSR6x34	371	275 - 468	mg/dL
			133	98.4 - 168	mmol/L
CA Calcium	Arsenazo III	OSR6x117	2.09	1.73 - 2.45	mmol/L
			8.37	6.95 - 9.79	mg/dL
	o-Cresolphthalein Complexone	OSR6x13	*	-	mmol/L
				-	mg/dL
CL Chloride	ISE		58.0	49.9 - 66.1	mmol/L
			206	177 - 234	mg/dL
CREA Creatinine	Enzymatic	OSR61204	*	-	mg/dL
				-	mmol/L
	Modified Jaffe	OSR6x78	59.8	45.4 - 74.2	mg/dL
			5.29	4.02 - 6.56	mmol/L
GLUC Glucose	Hexokinase	OSR6x21	25.4	19.8 - 31.0	mg/dL
			1.41	1.10 - 1.72	mmol/L
MG Magnesium	Xylidyl Blue	OSR6189	2.32	1.72 - 2.92	mmol/L
			5.63	4.17 - 7.09	mg/dL
PO4 Phosphorus	Phosphomolybdate/UV	OSR6x22	24.2	18.9 - 29.5	mg/dL
			7.82	6.10 - 9.54	mmol/L
K Potassium	ISE		32.4	26.9 - 37.9	mmol/L
			127	105 - 148	mg/dL
Na Sodium	ISE		84.6	73.6 - 95.6	mmol/L
			194	169 - 220	mg/dL
TP Total Protein	Pyrogallol Red		8.72	6.63 - 10.8	mg/dL
			87.2	66.3 - 108	mg/L
UREA Urea	Urease/GLDH	OSR6x34	795	588 - 1002	mg/dL
			133	98.4 - 168	mmol/L
UA Uric Acid	Uricase	OSR6x98	7.62	5.64 - 9.60	mg/dL
			453	335 - 571	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

BECKMAN COULTER DXC/SNYCHRON

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB Microalbumin	Immunoturbidimetric	MA	*	-	mg/L
				-	mg/dL
AMY Amylase	Enzymatic	AMY	*	-	U/L
				-	µkat/L
	Enzymatic	AMY7	*	-	U/L
				-	µkat/L
BUN Urea N	Urease/GLDH	BUN	*	-	mg/dL
				-	mmol/L
CA Calcium	ISE Indirect	CALC	*	-	mmol/L
				-	mg/dL
CL Chloride	ISE Indirect	Cl	*	-	mmol/L
				-	mg/dL
CREA Creatinine	Enzymatic	CR-E	*	-	mg/dL
				-	mmol/L
	Jaffe	CREA / CR-S CREm	*	-	mg/dL
				-	mmol/L
GLUC Glucose	Hexokinase	GLU / GLUH	*	-	mg/dL
				-	mmol/L
	Glucose Oxidase	GLUCm	*	-	mg/dL
				-	mmol/L
MG Magnesium	Calmagite	MG	*	-	mmol/L
				-	mg/dL
PO4 Phosphorus	Phosphomolybdate -UV	PHS	*	-	mg/dL
				-	mmol/L
	Phosphomolybdate -UV	PHOSm	*	-	mg/dL
				-	mmol/L
K Potassium	ISE indirect	K	*	-	mmol/L
				-	mg/dL
Na Sodium	ISE indirect	NA	*	-	mmol/L
				-	mg/dL
TP Total Protein	Timed Endpoint	M-TP	*	-	mg/dL
				-	mg/L
UREA Urea	Urease/GLDH	UREA	*	-	mg/dL
				-	mmol/L
UA Uric Acid	Uricase	URIC	*	-	mg/dL
				-	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT **38739**

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

ORTHO-VITROS CHEMISTRY SYSTEMS

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		12.2	8.78 - 15.6	mg/L
Microalbumin			1.22	0.878 - 1.56	mg/dL
	Immunoturbidimetric (DiAgam)		*	-	mg/L
				-	mg/dL
AMY	Amylopectin, Colorimetric		42.7	34.2 - 51.2	U/L
Amylase			0.712	0.569 - 0.854	µkat/L
BUN	Urease		387	286 - 487	mg/dL
Urea N			138	102 - 174	mmol/L
CA	Arsenazo III		2.06	1.71 - 2.41	mmol/L
Calcium			8.27	6.86 - 9.68	mg/dL
CL	ISE direct		50.4	43.3 - 57.5	mmol/L
Chloride			179	154 - 204	mg/dL
CREA	Colorimetric/Enzymatic		71.4	54.3 - 88.5	mg/dL
Creatinine			6.31	4.80 - 7.83	mmol/L
GLUC	Glucose Oxidase		24.6	19.2 - 30.0	mg/dL
Glucose			1.37	1.06 - 1.67	mmol/L
MG	Formazan Dye		2.83	2.09 - 3.57	mmol/L
Magnesium			6.88	5.09 - 8.67	mg/dL
PO4	Phosphomolybdate		27.7	21.6 - 33.8	mg/dL
Phosphorus			8.94	6.98 - 10.9	mmol/L
K	ISE direct		33.5	27.8 - 39.2	mmol/L
Potassium			131	109 - 153	mg/dL
Na	ISE direct		90.6	78.8 - 102	mmol/L
Sodium			208	181 - 235	mg/dL
TP	Biuret (UPRO)		21.3	16.2 - 26.4	mg/dL
Total Protein			213	162 - 264	mg/L
UREA	Urease		828	613 - 1043	mg/dL
Urea			138	102 - 174	mmol/L
UA	Uricase		*	-	mg/dL
Uric Acid				-	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

ROCHE COBAS C / E / MODULAR

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		22.6	16.3 - 28.9	mg/L
Microalbumin			2.26	1.63 - 2.89	mg/dL
AMY	Colorimetric	AMYL2	49.3	39.4 - 59.2	U/L
Amylase			0.822	0.657 - 0.986	µkat/L
BUN	Urease/GLDH	U-BUN	360	266 - 454	mg/dL
Urea N		UBUNU	128	94.7 - 161	mmol/L
CA	o-Cresolphthalein Complexone	CA	*	-	mmol/L
Calcium				-	mg/dL
	NM-Bapta	CA2	2.12	1.76 - 2.48	mmol/L
			8.51	7.06 - 9.96	mg/dL
CL	ISE indirect		48.4	41.6 - 55.2	mmol/L
Chloride			172	148 - 196	mg/dL
CREA	Jaffe	CREJ2	70.2	53.3 - 87.0	mg/dL
Creatinine			6.20	4.71 - 7.69	mmol/L
	Enzymatic	CREP2	75.4	57.3 - 93.5	mg/dL
			6.67	5.07 - 8.27	mmol/L
GLUC	Hexokinase	GLUC2	24.6	19.2 - 30.0	mg/dL
Glucose			1.37	1.06 - 1.67	mmol/L
		GLUC3	23.6	18.4 - 28.8	mg/dL
			1.31	1.02 - 1.60	mmol/L
MG	Chlorophosphonazo III	MG	*	-	mmol/L
Magnesium				-	mg/dL
	Xylidyl Blue	MG2	2.42	1.79 - 3.05	mmol/L
			5.89	4.36 - 7.42	mg/dL
PO4	Phosphomolybdate UV	PHOS2	24.5	19.1 - 29.9	mg/dL
Phosphorus			7.92	6.18 - 9.66	mmol/L
K	ISE indirect		31.8	26.4 - 37.2	mmol/L
Potassium			124	103 - 145	mg/dL
Na	ISE indirect		84.2	73.3 - 95.1	mmol/L
Sodium			194	168 - 219	mg/dL
TP	Turbidimetric	TPUC3	6.43	4.89 - 7.97	mg/dL
Total Protein			64.3	48.9 - 79.7	mg/L
UREA	Urease/GLDH	UREAL	776	574 - 978	mg/dL
Urea		URELU	129	95.6 - 163	mmol/L
UA	Uricase/Colorimetric	UA2	5.85	4.33 - 7.37	mg/dL
Uric Acid			348	257 - 438	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

ROCHE INTEGRA

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB Microalbumin	Immunoturbidimetric		*	-	mg/L
				-	mg/dL
AMY Amylase	EPS-G7PNP (Alpha)		*	-	U/L
				-	µkat/L
BUN Urea N	Urease		*	-	mg/dL
				-	mmol/L
CA Calcium	o-Cresolphthalein Complexone		*	-	mmol/L
				-	mg/dL
	NM-Bapta (2nd Gen.)		*	-	mmol/L
				-	mg/dL
CL Chloride	ISE indirect		*	-	mmol/L
				-	mg/dL
CREA Creatinine	Kinetic Alkaline Picrate (Jaffe)		*	-	mg/dL
				-	mmol/L
	Enzymatic/PAP		*	-	mg/dL
				-	mmol/L
GLUC Glucose	Hexokinase		*	-	mg/dL
				-	mmol/L
MG Magnesium	Chlorophosphonazo III		*	-	mmol/L
				-	mg/dL
PO4 Phosphorus	Phosphomolybdate-UV		*	-	mg/dL
				-	mmol/L
K Potassium	ISE indirect		*	-	mmol/L
				-	mg/dL
Na Sodium	ISE indirect		*	-	mmol/L
				-	mg/dL
TP Total Protein	Benzathonium Chloride		*	-	mg/dL
				-	mg/L
UREA Urea	Urease/GLDH		*	-	mg/dL
				-	mmol/L
UA Uric Acid	Uricase		*	-	mg/dL
				-	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

SIEMENS ADVIA

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	PEG-enhanced Immunoturbidimetric	µALB_2	*	-	mg/L
Microalbumin				-	mg/dL
AMY	G7 PNP	AMYLAS	*	-	U/L
Amylase				-	µkat/L
BUN	Uricase/Peroxidase	UA	*	-	mg/dL
Urea N		UA_c		-	mmol/L
CA	Arsenazo III	CA_2	*	-	mmol/L
Calcium		CA_2c		-	mg/dL
	CPC	CA	*	-	mmol/L
	for Advia 1200	CA_c		-	mg/dL
	CPC	CA	*	-	mmol/L
	for Advia 1650/1800 /2400	CA_c		-	mg/dL
CL	ISE indirect	Cl	*	-	mmol/L
Chloride				-	mg/dL
CREA	Jaffe	CREA_2	*	-	mg/dL
Creatinine	for Advia 1200	CRE_2c		-	mmol/L
	Enzymatic/PAP	ECRE_2	*	-	mg/dL
				-	mmol/L
GLUC	Hexokinase	GLUH_c	*	-	mg/dL
Glucose		GLUH_3		-	mmol/L
	Glucose Oxidase Trinder	GLUO	*	-	mg/dL
		GLUO_c		-	mmol/L
MG	Xylidyl Blue	MG	*	-	mmol/L
Magnesium				-	mg/dL
	Xylidyl Blue	MG_c	*	-	mmol/L
				-	mg/dL
PO4	Phosphomolybdate/UV	IP	*	-	mg/dL
Phosphorus				-	mmol/L
	Phosphomolybdate/UV	IP_c	*	-	mg/dL
				-	mmol/L
K	ISE indirect		*	-	mmol/L
Potassium				-	mg/dL
Na	ISE indirect		*	-	mmol/L
Sodium				-	mg/dL
TP	Dye Binding	UPRO_2	*	-	mg/dL
Total Protein				-	mg/L
UREA	Urease w/ GLDH	UN	*	-	mg/dL
Urea		UN_c		-	mmol/L
UA	Uricase/Peroxidase	UA	*	-	mg/dL
Uric Acid		UA_c		-	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

SIEMENS ATELLICA

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	PEG-enhanced Immunoturbidimetric	µALB_2	10.0	7.20 - 12.8	mg/L
Microalbumin			1.00	0.720 - 1.28	mg/dL
AMY	Ed-G7PNP	AMY_2	50.7	40.6 - 60.8	U/L
Amylase			0.845	0.676 - 1.01	µkat/L
BUN	Urease/GLDH	UN_c	363	269 - 457	mg/dL
Urea N			129	95.7 - 163	mmol/L
CA	Arsenazo III	CA_2	2.06	1.71 - 2.41	mmol/L
Calcium			8.25	6.85 - 9.65	mg/dL
	o-Cresolphthalein Complexone	Ca	2.06	1.71 - 2.41	mmol/L
			8.25	6.85 - 9.66	mg/dL
CL	Indirect IMT		66.9	57.5 - 76.3	mmol/L
Chloride			237	204 - 270	mg/dL
CORT	Chemiluminescence	Cor	*	-	µg/dL
Cortisol				-	nmol/L
CREA	Alkaline Picrate	Crea_2	66.2	50.3 - 82.1	mg/dL
Creatinine			5.85	4.45 - 7.26	mmol/L
	Enzymatic	ECre_2	72.7	55.2 - 90.1	mg/dL
		ECre3	6.42	4.88 - 7.96	mmol/L
GLUC	Glucose Oxidase	GluO	*	-	mg/dL
Glucose				-	mmol/L
	Hexokinase	GluH_3	25.3	19.8 - 30.9	mg/dL
			1.41	1.10 - 1.72	mmol/L
MG	Xylidyl Blue	Mg	2.40	1.78 - 3.03	mmol/L
Magnesium			5.84	4.32 - 7.36	mg/dL
PO4	Phosphomolybdate / UV	IP	25.0	19.5 - 30.5	mg/dL
Phosphorus			8.06	6.29 - 9.83	mmol/L
K	Indirect IMT		32.3	26.8 - 37.8	mmol/L
Potassium			126	105 - 148	mg/dL
Na	Indirect IMT		83.9	73.0 - 94.8	mmol/L
Sodium			193	168 - 218	mg/dL
TP	Pyrogallol Red-Molybdate	UCFP	13.9	10.6 - 17.2	mg/dL
Total Protein			139	106 - 172	mg/L
UREA	Urease/GLDH	UN_c	777	575 - 979	mg/dL
Urea			129	95.7 - 163	mmol/L
UA	Uricase / Peroxidase	UA	6.49	4.80 - 8.18	mg/dL
Uric Acid			386	286 - 486	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

CE IVD

SIEMENS BNII/PROSPEC

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Nephelometric		*	-	mg/L
Microalbumin				-	mg/dL
TP	Nephelometric		*	-	mg/dL
Total Protein				-	mg/L

* not available at time of printing

SIEMENS DCA 2000

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		*	-	mg/L
Microalbumin				-	mg/dL
CREA	3,5 dinitrobenzoic acid		*	-	mg/dL
Creatinine				-	mmol/L

BECKMAN COULTER IMMAGE

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Rate Nephelometric	MA	*	-	mg/L
Microalbumin				-	mg/dL

Flame Photometer

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
K	Flame Photometry		*	-	mmol/L
Potassium				-	mg/dL
Na	Flame Photometry		*	-	mmol/L
Sodium				-	mg/dL

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

SIEMENS DIMENSION

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	PETINIA	MALB	11.3	8.14 - 14.5	mg/L
Microalbumin			1.13	0.814 - 1.45	mg/dL
AMY	CNPG3	AMY	38.9	31.1 - 46.7	U/L
Amylase			0.648	0.519 - 0.778	µkat/L
BUN	Urease/GLDH		394	292 - 496	mg/dL
Urea N			141	104 - 178	mmol/L
CA	o-Cresolphthalein Complexone	CA	2.21	1.83 - 2.59	mmol/L
Calcium			8.82	7.32 - 10.3	mg/dL
CL	Indirect IMT	RxL	77.8	66.9 - 88.7	mmol/L
Chloride			276	237 - 315	mg/dL
	Indirect IMT	Xpand	80.8	69.5 - 92.1	mmol/L
		EXL	287	246 - 327	mg/dL
CREA	Modified Jaffe	CRE2	71.6	54.4 - 88.8	mg/dL
Creatinine			6.33	4.81 - 7.85	mmol/L
	Enzymatic	EZCR	*	-	mg/dL
				-	mmol/L
GLUC	Hexokinase	GLUC	*	-	mg/dL
Glucose				-	mmol/L
MG	Methylthymol Blue	MG	2.54	1.88 - 3.20	mmol/L
Magnesium			6.17	4.57 - 7.78	mg/dL
PO4	Phosphomolybdate	PHOS	23.1	18.0 - 28.2	mg/dL
Phosphorus			7.47	5.83 - 9.11	mmol/L
K	Indirect IMT	RxL	30.0	24.9 - 35.1	mmol/L
Potassium			117	97.4 - 137	mg/dL
	Indirect IMT	Xpand	32.2	26.7 - 37.7	mmol/L
		EXL	126	104 - 147	mg/dL
Na	Indirect IMT	RxL	80.8	70.3 - 91.3	mmol/L
Sodium			186	162 - 210	mg/dL
	Indirect IMT	Xpand	83.8	72.9 - 94.7	mmol/L
		EXL	193	168 - 218	mg/dL
TP	Pyrogallol Red	UCFP	11.4	8.66 - 14.1	mg/dL
Total Protein			114	86.6 - 141	mg/L
UREA	Urease/GLDH		842	623 - 1061	mg/dL
Urea			141	104 - 178	mmol/L
UA	Uricase	URCA	6.06	4.48 - 7.64	mg/dL
Uric Acid			360	267 - 454	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

SIEMENS DIMENSION VISTA

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Nephelometric	MALB	34.4	24.8 - 44.0	mg/L
Microalbumin			3.44	2.48 - 4.40	mg/dL
AMY	CNPG3	AMY	41.8	33.4 - 50.2	U/L
Amylase			0.697	0.557 - 0.836	µkat/L
BUN	Urease/GLDH		362	268 - 456	mg/dL
Urea N			129	95.5 - 163	mmol/L
CA	o-Cresolphthalein Complexone	CA	2.15	1.78 - 2.52	mmol/L
Calcium			8.59	7.13 - 10.1	mg/dL
CL	Indirect IMT	CL	60.2	51.8 - 68.6	mmol/L
Chloride			213	184 - 243	mg/dL
CREA	Modified Jaffe	CRE2	74.1	56.3 - 91.9	mg/dL
Creatinine			6.55	4.98 - 8.12	mmol/L
	Enzymatic	ECREA	72.0	54.7 - 89.3	mg/dL
			6.37	4.84 - 7.90	mmol/L
GLUC	Hexokinase	GLU	24.4	19.0 - 29.8	mg/dL
Glucose			1.35	1.06 - 1.65	mmol/L
MG	Methylthymol Blue	MG	2.44	1.81 - 3.07	mmol/L
Magnesium			5.94	4.40 - 7.48	mg/dL
PO4	Phosphomolybdate	PHOS	23.3	18.2 - 28.4	mg/dL
Phosphorus			7.52	5.87 - 9.18	mmol/L
K	Indirect IMT	K	31.8	26.4 - 37.2	mmol/L
Potassium			124	103 - 145	mg/dL
Na	Indirect IMT	NA	72.8	63.3 - 82.3	mmol/L
Sodium			167	146 - 189	mg/dL
TP	Pyrogallol Red	UCFP	13.8	10.5 - 17.1	mg/dL
Total Protein			138	105 - 171	mg/L
UREA	Urease/GLDH		775	574 - 977	mg/dL
Urea			129	95.5 - 163	mmol/L
UA	Uricase	URCA	8.46	6.26 - 10.7	mg/dL
Uric Acid			503	372 - 634	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

MINDRAY BS SERIES

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		23.8	17.1 - 30.5	mg/L
Microalbumin			2.38	1.71 - 3.05	mg/dL
	Immunoturbidimetric (Rgt lot # 46025004)		18.8	13.5 - 24.1	mg/L
			1.88	1.35 - 2.41	mg/dL
AMY	Ed-G7PNP, IFCC		56.0	44.8 - 67.2	U/L
Amylase			0.934	0.747 - 1.12	µkat/L
BUN	Urease/GLDH, UV		389	288 - 490	mg/dL
Urea N			138	102 - 174	mmol/L
CA	Arsenazo III		2.01	1.67 - 2.35	mmol/L
Calcium			8.02	6.66 - 9.38	mg/dL
CL	ISE indirect		62.0	53.3 - 70.7	mmol/L
Chloride			220	189 - 251	mg/dL
CORT	CLIA		3.32	2.66 - 3.98	µg/dL
Cortisol			91.5	73.2 - 110	nmol/L
CREA	Sarcosine Oxidase		73.9	56.2 - 91.7	mg/dL
Creatinine			6.54	4.97 - 8.11	mmol/L
GLUC	Hexokinase		26.1	20.4 - 31.8	mg/dL
Glucose			1.45	1.13 - 1.77	mmol/L
MG	Xylidyl Blue (Mg II)		2.44	1.80 - 3.07	mmol/L
Magnesium			5.92	4.38 - 7.46	mg/dL
PO4	Phosphomolybdate (P II)		23.6	18.4 - 28.8	mg/dL
Phosphorus			7.63	5.95 - 9.31	mmol/L
K	ISE indirect		32.3	26.8 - 37.8	mmol/L
Potassium			126	105 - 148	mg/dL
Na	ISE indirect		84.2	73.3 - 95.1	mmol/L
Sodium			194	168 - 219	mg/dL
TP	Pyrogallol Red-Molybdate (TPUC)		6.42	4.88 - 7.96	mg/dL
Total Protein			64.2	48.8 - 79.6	mg/L
UREA	Urease/GLDH, UV		834	617 - 1050	mg/dL
Urea			138	102 - 174	mmol/L
UA	Uricase/Peroxidase		6.29	4.65 - 7.93	mg/dL
Uric Acid			374	277 - 471	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

THERMO FISHER SCIENTIFIC INDIKO SYSTEMS

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		34.7	25.0 - 44.4	mg/L
Microalbumin			3.47	2.50 - 4.44	mg/dL
AMY	IFCC		49.2	39.4 - 59.0	U/L
Amylase			0.820	0.656 - 0.98	µkat/L
BUN	Urease/GLDH		411	304 - 518	mg/dL
Urea N			147	109 - 185	mmol/L
CA	Arsenazo III		1.99	1.65 - 2.33	mmol/L
Calcium			7.97	6.62 - 9.32	mg/dL
CREA	Jaffé		69.8	53.1 - 86.6	mg/dL
Creatinine			6.18	4.70 - 7.66	mmol/L
GLUC	Hexokinase		25.3	19.8 - 30.9	mg/dL
Glucose			1.41	1.10 - 1.72	mmol/L
PO4	Molybdate, UV		24.7	19.3 - 30.1	mg/dL
Phosphorus			7.98	6.22 - 9.73	mmol/L
TP	Pyrogallol Red-Molybdate		< 8.00	-	mg/dL
Total Protein			< 80.0	-	mg/L
UREA	Urease/GLDH		880	651 - 1109	mg/dL
Urea			147	109 - 185	mmol/L
UA	Uricase		6.79	5.02 - 8.56	mg/dL
Uric Acid			404	299 - 509	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 1

LOT 38739

Cat. No. 5 9170 99 10 061 (6 x 5 mL)
5 9170 99 10 062 (20 x 5 mL)

2027-04-30

THERMO FISHER SCIENTIFIC KONELAB SYSTEMS

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CREA	Jaffé		62.2	47.3 - 77.1	mg/dL
Creatinine			5.51	4.19 - 6.83	mmol/L
GLUC	Hexokinase		24.6	19.2 - 30.0	mg/dL
Glucose			1.37	1.06 - 1.67	mmol/L
TP	Pyrogallol Red-Molybdate		< 8.00	-	mg/dL
Total Protein			< 80.0	-	mg/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com