

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 **2027-04-30**



Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|--|-------------------------------|------------------|--------------|----------------------------|-----------------|
| ALB Microalbumin | Immunoturbidimetric | 1 0242 | 55.5 5.55 | 40.0 - 71.1 4.00 - 7.11 | mg/L mg/dL |
| AMY Amylase | EPS G-7 | 1 0501 | 213 3.54 | 170 - 255 2.84 - 4.25 | U/L µkat/L |
| BUN Urea N | Urease UV | 1 3101 | 793 283 | 587 - 1000 209 - 356 | mg/dL mmol/L |
| CAAS Calcium | Arsenazo | 1 1130 | 2.66 10.7 | 2.21 - 3.12 8.87 - 12.5 | mmol/L mg/dL |
| CA Calcium | Phosphonazo | 1 1181 | 2.76 11.1 | 2.29 - 3.23 9.18 - 12.9 | mmol/L mg/dL |
| CREA Creatinine | Jaffé, without compensation | 1 1711 | 146 12.9 | 111 - 181 9.78 - 16.0 | mg/dL mmol/L |
| CREAP Creatinine | Enzymatic PAP | 1 1759 | 154 13.6 | 117 - 191 10.3 - 16.9 | mg/dL mmol/L |
| D Specific Gravity | Refractrometer | | 1.021 | 1.010 - 1.035 | |
| GLUHK Glucose | HK/G6P-DH | 1 2511 | 269 14.9 | 210 - 328 11.6 - 18.2 | mg/dL mmol/L |
| hCG Human Chorionic Gonadotropin | | | | Positive | |
| MG Magnesium | Xylidyl blue | 1 4610 | 5.03 12.2 | 3.72 - 6.33 9.04 - 15.4 | mmol/L mg/dL |
| OSMO Osmolality | Freezing point | | 800 | 640 - 960 | mOsmo/kg |
| pH | pH Meter | | 6.21 | 5.50 - 6.90 | |
| PO4 Phosphorus | Molybdate UV, substrate start | 1 5211 | 49.0 15.8 | 38.2 - 59.7 12.3 - 19.3 | mg/dL mmol/L |
| TP Total Protein | Pyrogallol red | 1 0210 | 60.1 601 | 45.7 - 74.5 457 - 745 | mg/dL mg/L |
| UREA Urea | Urease UV | 1 3101 | 1699 283 | 1257 - 2141 209 - 356 | mg/dL mmol/L |
| UA TBH Uric Acid | TBHBA | 1 3021 | 13.1 778 | 9.68 - 16.5 576 - 980 | mg/dL µmol/L |
| UA Uric Acid | TOOS | 1 3001 | 13.2 786 | 9.78 - 16.7 582 - 991 | mg/dL µmol/L |


* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

respons[®]910



Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|------------------------------|------------------|--------------|----------------------------|-----------------|
| ALB Microalbumin | Immunoturbidimetric | 1 0242 | 55.5 5.55 | 40.0 - 71.1 4.00 - 7.11 | mg/L mg/dL |
| AMY Amylase | EPS G-7 | 1 0501 | 213 3.54 | 170 - 255 2.84 - 4.25 | U/L µkat/L |
| BUN Urea N | Urease UV | 1 3101 | 793 283 | 587 - 1000 209 - 356 | mg/dL mmol/L |
| CA Calcium | Phosphonazo | 1 1181 | 2.76 11.1 | 2.29 - 3.23 9.18 - 12.9 | mmol/L mg/dL |
| CREA Creatinine | Jaffé , without compensation | 1 1711 | 146 12.9 | 111 - 181 9.78 - 16.0 | mg/dL mmol/L |
| CREAP Creatinine | Enzymatic PAP | 1 1759 | 154 13.6 | 117 - 191 10.3 - 16.9 | mg/dL mmol/L |
| GLUHK Glucose | HK/G6P-DH | 1 2511 | 269 14.9 | 210 - 328 11.6 - 18.2 | mg/dL mmol/L |
| MG Magnesium | Xylidyl blue | 1 4610 | 5.03 12.2 | 3.72 - 6.33 9.04 - 15.4 | mmol/L mg/dL |
| PO4 Phosphorus | Molybdate UV | 1 5211 | 49.0 15.8 | 38.2 - 59.7 12.3 - 19.3 | mg/dL mmol/L |
| TP Total Protein | Pyrogallol red | 1 0210 | 60.1 601 | 45.7 - 74.5 457 - 745 | mg/dL mg/L |
| UREA Urea | Urease UV | 1 3101 | 1699 283 | 1257 - 2141 209 - 356 | mg/dL mmol/L |
| UA Uric Acid | TOOS | 1 3001 | 13.2 786 | 9.78 - 16.7 582 - 991 | mg/dL µmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

respons[®]920



Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | 1 SD | Unit |
|----------------------------|------------------------------|------------------|--------------|----------------------------|----------------|-----------------|
| ALB Microalbumin | Immunoturbidimetric | 1 0242 | 55.5 5.55 | 40.0 - 71.1 4.00 - 7.11 | 5.18 0.518 | mg/L mg/dL |
| AMY Amylase | EPS G-7 | 1 0501 | 213 3.54 | 170 - 255 2.84 - 4.25 | 14.2 0.236 | U/L µkat/L |
| BUN Urea N | Urease UV | 1 3101 | 793 283 | 587 - 1000 209 - 356 | 68.8 24.5 | mg/dL mmol/L |
| CA Calcium | Phosphonazo | 1 1181 | 2.76 11.1 | 2.29 - 3.23 9.18 - 12.9 | 0.156 0.627 | mmol/L mg/dL |
| CL Chloride | Direct ISE | 960405 | 146 516 | 125 - 166 444 - 588 | 6.79 24.1 | mmol/L mg/dL |
| CREA Creatinine | Jaffé , without compensation | 1 1711 | 146 12.9 | 111 - 181 9.78 - 16.0 | 11.7 1.03 | mg/dL mmol/L |
| CREAP Creatinine | Enzymatic PAP | 1 1759 | 154 13.6 | 117 - 191 10.3 - 16.9 | 12.3 1.09 | mg/dL µmol/L |
| GLUHK Glucose | HK/G6P-DH | 1 2511 | 269 14.9 | 210 - 328 11.6 - 18.2 | 19.7 1.09 | mg/dL mmol/L |
| MG Magnesium | Xylidyl blue | 1 4610 | 5.03 12.2 | 3.72 - 6.33 9.04 - 15.4 | 0.436 1.06 | mmol/L mg/dL |
| PO4 Phosphorus | Molybdate UV | 1 5211 | 49.0 15.8 | 38.2 - 59.7 12.3 - 19.3 | 3.59 1.16 | mg/dL mmol/L |
| K Potassium | Direct ISE | 960403 | 77.8 304 | 64.5 - 91.0 252 - 356 | 4.41 17.2 | mmol/L mg/dL |
| Na Sodium | Direct ISE | 960402 | 163 376 | 142 - 185 327 - 424 | 7.08 16.3 | mmol/L mg/dL |
| UREA Urea | Urease UV | 1 3101 | 1699 283 | 1257 - 2141 209 - 356 | 147 24.5 | mg/dL mmol/L |
| UA Uric Acid | TOOS | 1 3001 | 13.2 786 | 9.78 - 16.7 582 - 991 | 1.15 68.1 | mg/dL µmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

respons[®]940



Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | 1 SD | Unit |
|----------------------------|------------------------------|------------------|--------------|----------------------------|----------------|-----------------|
| ALB Microalbumin | Immunoturbidimetric | 1 0242 | 55.5 5.55 | 40.0 - 71.1 4.00 - 7.11 | 5.18 0.518 | mg/L mg/dL |
| AMY Amylase | EPS G-7 | 1 0501 | 213 3.54 | 170 - 255 2.84 - 4.25 | 14.2 0.236 | U/L µkat/L |
| BUN Urea N | Urease UV | 1 3101 | 793 283 | 587 - 1000 209 - 356 | 68.8 24.5 | mg/dL mmol/L |
| CA Calcium | Phosphonazo | 1 1181 | 2.76 11.1 | 2.29 - 3.23 9.18 - 12.9 | 0.156 0.627 | mmol/L mg/dL |
| CL Chloride | Direct ISE | 960405 | 146 516 | 125 - 166 444 - 588 | 6.79 24.1 | mmol/L mg/dL |
| CREA Creatinine | Jaffé , without compensation | 1 1711 | 146 12.9 | 111 - 181 9.78 - 16.0 | 11.7 1.03 | mg/dL mmol/L |
| CREAP Creatinine | Enzymatic PAP | 1 1759 | 154 13.6 | 117 - 191 10.3 - 16.9 | 12.3 1.09 | mg/dL mmol/L |
| GLUHK Glucose | HK/G6P-DH | 1 2511 | 269 14.9 | 210 - 328 11.6 - 18.2 | 19.7 1.09 | mg/dL mmol/L |
| MG Magnesium | Xylidyl blue | 1 4610 | 5.03 12.2 | 3.72 - 6.33 9.04 - 15.4 | 0.436 1.06 | mmol/L mg/dL |
| PO4 Phosphorus | Molybdate UV | 1 5211 | 49.0 15.8 | 38.2 - 59.7 12.3 - 19.3 | 3.59 1.16 | mg/dL mmol/L |
| K Potassium | Direct ISE | 960403 | 77.8 304 | 64.5 - 91.0 252 - 356 | 4.41 17.2 | mmol/L mg/dL |
| Na Sodium | Direct ISE | 960402 | 163 376 | 142 - 185 327 - 424 | 7.08 16.3 | mmol/L mg/dL |
| TP Total Protein | Pyrogallol red | 1 0210 | 60.1 601 | 45.7 - 74.5 457 - 745 | 4.81 48.1 | mg/dL mg/L |
| UREA Urea | Urease UV | 1 3101 | 1699 283 | 1257 - 2141 209 - 356 | 147 24.5 | mg/dL mmol/L |
| UA Uric Acid | TOOS | 1 3001 | 13.2 786 | 9.78 - 16.7 582 - 991 | 1.15 68.1 | mg/dL µmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

BioMajesty[®]

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | 1 SD | Unit |
|----------------------------|-----------------------------|------------------|--------------|----------------------------|----------------|-----------------|
| ALB Microalbumin | Immunoturbidimetric | 1 0242 | 55.5 5.55 | 40.0 - 71.1 4.00 - 7.11 | 5.18 0.518 | mg/L mg/dL |
| AMY Amylase | EPS G-7 | 1 0501 | 213 3.54 | 170 - 255 2.84 - 4.25 | 14.2 0.236 | U/L µkat/L |
| BUN Urea N | Urease UV | 1 3101 | 793 283 | 587 - 1000 209 - 356 | 68.8 24.5 | mg/dL mmol/L |
| CA Calcium | Phosphonazo | 1 1181 | 2.76 11.1 | 2.29 - 3.23 9.18 - 12.9 | 0.156 0.627 | mmol/L mg/dL |
| CL Chloride | Indirect ISE | 980132 | 168 596 | 144 - 191 512 - 679 | 7.84 27.8 | mmol/L mg/dL |
| CREA Creatinine | Jaffé, without compensation | 1 1711 | 146 12.9 | 111 - 181 9.78 - 16.0 | 11.7 1.03 | mg/dL mmol/L |
| CREAP Creatinine | Enzymatic PAP | 1 1759 | 154 13.6 | 117 - 191 10.3 - 16.9 | 12.3 1.09 | mg/dL mmol/L |
| GLUHK Glucose | HK/G6P-DH | 1 2511 | 269 14.9 | 210 - 328 11.6 - 18.2 | 19.7 1.09 | mg/dL mmol/L |
| MG Magnesium | Xylidyl blue | 1 4610 | 5.03 12.2 | 3.72 - 6.33 9.04 - 15.4 | 0.436 1.06 | mmol/L mg/dL |
| PO4 Phosphorus | Molybdate UV | 1 5211 | 49.0 15.8 | 38.2 - 59.7 12.3 - 19.3 | 3.59 1.16 | mg/dL mmol/L |
| K Potassium | Indirect ISE | 980131 | 85.0 332 | 70.5 - 99.4 276 - 389 | 4.81 18.8 | mmol/L mg/dL |
| Na Sodium | Indirect ISE | 980130 | 184 423 | 160 - 208 368 - 478 | 7.98 18.3 | mmol/L mg/dL |
| TP Total Protein | Pyrogallol red | 1 0210 | 60.1 601 | 45.7 - 74.5 457 - 745 | 4.81 48.1 | mg/dL mg/L |
| UREA Urea | Urease UV | 1 3101 | 1699 283 | 1257 - 2141 209 - 356 | 147 24.5 | mg/dL mmol/L |
| UA Uric Acid | TOOS | 1 3001 | 13.2 786 | 9.78 - 16.7 582 - 991 | 1.15 68.1 | mg/dL µmol/L |


* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

ABBOTT AEROSSET/ARCHITECT



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|-------------------------------------|------------------|-------|-------------|-------------|
| ALB Microalbumin | Turbidimetric / Immunoturbidimetric | 2K98 | 94.9 | 68.3 - 121 | mg/L |
| | | | 9.49 | 6.83 - 12.1 | mg/dL |
| AMY Amylase | CNPG3 37° C | 7D58 | * | - | U/L |
| | CNPG3 37° C, IFCC | 7D58 | * | - | U/L |
| | | | - | - | µkat/L |
| BUN Urea N | Urease/GLDH | 7D75 | * | - | mg/dL |
| | | | - | - | mmol/L |
| CA Calcium | Arsenazo III | 3L79 | 2.70 | 2.24 - 3.16 | mmol/L |
| | | | 10.8 | 8.99 - 12.7 | mg/dL |
| CL Chloride | ISE Indirect | 2P32 | 147 | 126 - 168 | mmol/L |
| | | | 521 | 448 - 594 | mg/dL |
| CORT Cortisol | CMIA | 8D15 | * | - | µg/dL |
| | | | - | - | nmol/L |
| CREA Creatinine | Alkaline Picrate | 3L81 | * | - | mg/dL |
| | | | - | - | mmol/L |
| | Enzymatic | 8L24 | 168 | 128 - 208 | mg/dL |
| GLUC Glucose | Hexokinase/G-6-PDH | 3L82 | 14.9 | 11.3 - 18.4 | mmol/L |
| | | | 280 | 219 - 342 | mg/dL |
| MG Magnesium | Arsenazo III | 7D70 | * | - | mmol/L |
| | | | - | - | mg/dL |
| PO4 Phosphorus | Phosphomolybdate | 7D71 | * | - | mg/dL |
| | | | - | - | mmol/L |
| | | | 4T07 | 43.8 | 34.2 - 53.4 |
| K Potassium | ISE Indirect | 2P32 | 14.1 | 11.0 - 17.3 | mmol/L |
| | | | 83.0 | 68.9 - 97.1 | mmol/L |
| Na Sodium | ISE Indirect | 2P32 | 325 | 269 - 380 | mg/dL |
| | | | 183 | 159 - 207 | mmol/L |
| TP Total Protein | Benzethonium Chloride | 7D79 | 421 | 366 - 475 | mg/dL |
| | | | 59.4 | 45.1 - 73.7 | mg/dL |
| UREA Urea | Urease/GLDH | 7D75 | 594 | 451 - 737 | mg/L |
| | | | * | - | mg/dL |
| | | | 4T12 | 1675 | 1240 - 2111 |
| UA Uric Acid | Uricase (UAU) | 3P39 | 279 | 206 - 351 | mmol/L |
| | | | * | - | mg/dL |
| | | | 4T13 | 12.5 | 9.25 - 15.8 |
| | | | 744 | 550 - 937 | µmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT

38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)



2027-04-30

ABBOTT ALINITY




Value sheet version: 1

| Short Name / Analyte | Method | Convers | Reagent Cat. No. | Value | Range | Unit | |
|----------------------------|-----------------------------|---------|------------------|-----------|-------------|-------------|-------------|
| ALB Microalbumin | Immunoturbidimetric | 0.1 | 08P04 | 93.6 | 67.4 - 120 | mg/L | |
| | | | | 9.36 | 6.74 - 12.0 | mg/dL | |
| AMY Amylase | Enzymatic CNPG3 37° C | 0.0167 | 07P58 | 233 | 186 - 280 | U/L | |
| | | | | 3.88 | 3.11 - 4.66 | µkat/L | |
| | | | | * | - | U/L | |
| | Enzymatic CNPG3 37° C, IFCC | 0.0167 | 07P58 | | - | µkat/L | |
| BUN Urea N | Urease/GLDH | 0.3563 | 08P16 | 709 | 525 - 893 | mg/dL | |
| | | | | 253 | 187 - 318 | mmol/L | |
| CA Calcium | Arsenazo III | 4.01 | 07P57 | 2.67 | 2.22 - 3.12 | mmol/L | |
| | | | | 10 | 8.89 - 12.5 | mg/dL | |
| CL Chloride | ISE Indirect | 3.5461 | 07P53 | 149 | 128 - 170 | mmol/L | |
| | | | | 10 | 528 | 454 - 602 | mg/dL |
| CORT Cortisol | CMIA | 27.59 | 08P33 | 18.7 | 15.0 - 22.4 | µg/dL | |
| | | | | 515 | 412 - 618 | nmol/L | |
| CREA Creatinine | Alkaline Picrate | 0.0884 | 07P99 | 166 | 126 - 206 | mg/dL | |
| | | | | 10 | 14.6 | 11.1 - 18.1 | mmol/L |
| | | | | Enzymatic | 0.0884 | 08P01 | 167 |
| | | | | 14.8 | 11.2 - 18.3 | mmol/L | |
| GLUC Glucose | Hexokinase/G-6-PDH | 0.0555 | 07P55 | 283 | 221 - 345 | mg/dL | |
| | | | | 0.01 | 15.7 | 12.3 - 19.2 | mmol/L |
| MG Magnesium | Enzymatic | 2.4307 | 08P19 | 5.35 | 3.96 - 6.74 | mmol/L | |
| | | | | 13.0 | 9.62 - 16.4 | mg/dL | |
| PO4 Phosphorus | Phosphomolybdate | 0.3229 | 04U03 | 42.1 | 32.8 - 51.4 | mg/dL | |
| | | | | 13.6 | 10.6 - 16.6 | mmol/L | |
| K Potassium | ISE Indirect | 3.9098 | 07P53 | 84.5 | 70.1 - 98.9 | mmol/L | |
| | | | | 330 | 274 - 387 | mg/dL | |
| Na Sodium | ISE Indirect | 2.299 | 07P53 | 185 | 161 - 209 | mmol/L | |
| | | | | 425 | 370 - 481 | mg/dL | |
| TP Total Protein | Benzethonium Chloride | 10 | 07P59 | 57.1 | 43.4 - 70.8 | mg/dL | |
| | | | | 571 | 434 - 708 | mg/L | |
| UREA Urea | Urease/GLDH | 0.1665 | 08P16 | 1517 | 1123 - 1911 | mg/dL | |
| | | | | 253 | 187 - 318 | mmol/L | |
| UA Uric Acid | Uricase | 59.48 | 08P56 | 13.0 | 9.62 - 16.4 | mg/dL | |
| | | | | 773 | 572 - 974 | µmol/L | |
| | | | | 59.48 | 04U09 | 12.4 | 9.18 - 15.6 |
| | | | | 738 | 546 - 929 | µmol/L | |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com



TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30



BECKMAN COULTER ACCESS

Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|-------------------------|-------------------|------------------|-------|-------|-----------------|
| CORT Cortisol | Chemiluminescence | Cort | * | - | µg/dL nmol/L |

BECKMAN COULTER DXI

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|-------------------------|-------------------|------------------|-------|-------|-----------------|
| CORT Cortisol | Chemiluminescence | Cort | * | - | µg/dL nmol/L |

ORTHO-VITROS IMMUNOASSAY SYSTEMS

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|-------------------------|-------------------|------------------|-------|-------|-----------------|
| CORT Cortisol | Chemiluminescence | | * | - | µg/dL nmol/L |

SIEMENS CENTAUR


| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|-------------------------|-------------------|------------------|-------|-------|-----------------|
| CORT Cortisol | Chemiluminescence | | * | - | µg/dL nmol/L |



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

BECKMAN COULTER AU



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|------------------------------|------------------|--------------|----------------------------|-----------------|
| ALB Microalbumin | Turbidimetric | OSR6167 | * | - | mg/L |
| | Turbidimetric | UALB | 106 10.6 | 76.3 - 136 7.63 - 13.6 | mg/L mg/dL |
| AMY Amylase | CNPG3 | OSR6x06 | 182 3.04 | 146 - 218 2.43 - 3.64 | U/L µkat/L |
| | G7PNP/IFCC | OSR6182 | * | - | U/L µkat/L |
| BUN Urea N | Urease/GLDH | OSR6x34 | 720 257 | 533 - 907 190 - 323 | mg/dL mmol/L |
| | Arsenazo III | OSR6x117 | 2.87 11.5 | 2.38 - 3.36 9.55 - 13.5 | mmol/L mg/dL |
| CA Calcium | o-Cresolphthalein Complexone | OSR6x13 | * | - | mmol/L mg/dL |
| | ISE | | 164 582 | 141 - 187 500 - 663 | mmol/L mg/dL |
| CL Chloride | ISE | | 164 582 | 141 - 187 500 - 663 | mmol/L mg/dL |
| | Enzymatic | OSR61204 | * | - | mg/dL mmol/L |
| CREA Creatinine | Modified Jaffe | OSR6x78 | 161 14.2 | 122 - 200 10.8 - 17.6 | mg/dL mmol/L |
| | Hexokinase | OSR6x21 | 283 15.7 | 221 - 345 12.3 - 19.2 | mg/dL mmol/L |
| MG Magnesium | Xylidyl Blue | OSR6189 | 5.09 12.4 | 3.77 - 6.41 9.16 - 15.6 | mmol/L mg/dL |
| | Phosphomolybdate/UV | OSR6x22 | 43.8 14.1 | 34.2 - 53.4 11.0 - 17.3 | mg/dL mmol/L |
| K Potassium | ISE | | 83.8 328 | 69.6 - 98.0 272 - 383 | mmol/L mg/dL |
| | ISE | | 184 423 | 160 - 208 368 - 478 | mmol/L mg/dL |
| Na Sodium | ISE | | 184 423 | 160 - 208 368 - 478 | mmol/L mg/dL |
| | Pyrogallol Red | OSR6x70 | 58.8 588 | 44.7 - 72.9 447 - 729 | mg/dL mg/L |
| TP Total Protein | Pyrogallol Red | OSR6x70 | 58.8 588 | 44.7 - 72.9 447 - 729 | mg/dL mg/L |
| | Urease/GLDH | OSR6x34 | 1541 257 | 1140 - 1942 190 - 323 | mg/dL mmol/L |
| UREA Urea | Urease/GLDH | OSR6x34 | 1541 257 | 1140 - 1942 190 - 323 | mg/dL mmol/L |
| | Uricase | OSR6x98 | 15.3 910 | 11.3 - 19.3 673 - 1147 | mg/dL µmol/L |
| UA Uric Acid | Uricase | OSR6x98 | 15.3 910 | 11.3 - 19.3 673 - 1147 | mg/dL µmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

BECKMAN COULTER DXC/SNYCHRON



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|----------------------|---------------------|-------|-------|-----------------|
| ALB Microalbumin | Immunoturbidimetric | MA | * | - | mg/L mg/dL |
| AMY Amylase | Enzymatic | AMY | * | - | U/L µkat/L |
| | Enzymatic | AMY7 | * | - | U/L µkat/L |
| BUN Urea N | Urease/GLDH | BUN | * | - | mg/dL mmol/L |
| CA Calcium | ISE Indirect | CALC | * | - | mmol/L mg/dL |
| CL Chloride | ISE Indirect | Cl | * | - | mmol/L mg/dL |
| CREA Creatinine | Enzymatic | CR-E | * | - | mg/dL mmol/L |
| | Jaffe | CREA / CR-S CREm | * | - | mg/dL mmol/L |
| GLUC Glucose | Hexokinase | GLU / GLUH | * | - | mg/dL mmol/L |
| | Glucose Oxidase | GLUCm | * | - | mg/dL mmol/L |
| MG Magnesium | Calmagite | MG | * | - | mmol/L mg/dL |
| PO4 Phosphorus | Phosphomolybdate -UV | PHS | * | - | mg/dL mmol/L |
| | Phosphomolybdate -UV | PHOSm | * | - | mg/dL mmol/L |
| K Potassium | ISE indirect | K | * | - | mmol/L mg/dL |
| Na Sodium | ISE indirect | NA | * | - | mmol/L mg/dL |
| TP Total Protein | Timed Endpoint | M-TP | * | - | mg/dL mg/L |
| UREA Urea | Urease/GLDH | UREA | * | - | mg/dL mmol/L |
| UA Uric Acid | Uricase | URIC | * | - | mg/dL µmol/L |


* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

ORTHO-VITROS CHEMISTRY SYSTEMS



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|------------------------------|------------------|-------|-------------|---------------|
| ALB Microalbumin | Immunoturbidimetric | | 67.9 | 48.9 - 86.9 | mg/L |
| | | | 6.79 | 4.89 - 8.69 | mg/dL |
| | Immunoturbidimetric (DiAgam) | | * | - | mg/L mg/dL |
| AMY Amylase | Amylopectin, Colorimetric | | 161 | 129 - 193 | U/L |
| | | | 2.70 | 2.16 - 3.22 | µkat/L |
| BUN Urea N | Urease | | 739 | 547 - 931 | mg/dL |
| | | | 263 | 195 - 331 | mmol/L |
| CA Calcium | Arsenazo III | | 2.77 | 2.30 - 3.24 | mmol/L |
| | | | 11.1 | 9.22 - 13.0 | mg/dL |
| CL Chloride | ISE direct | | 143 | 123 - 163 | mmol/L |
| | | | 507 | 436 - 578 | mg/dL |
| CREA Creatinine | Colorimetric/Enzymatic | | 170 | 129 - 211 | mg/dL |
| | | | 15.1 | 11.5 - 18.7 | mmol/L |
| GLUC Glucose | Glucose Oxidase | | 283 | 221 - 345 | mg/dL |
| | | | 15.7 | 12.3 - 19.2 | mmol/L |
| MG Magnesium | Formazan Dye | | 5.71 | 4.23 - 7.19 | mmol/L |
| | | | 13.9 | 10.3 - 17.5 | mg/dL |
| PO4 Phosphorus | Phosphomolybdate | | 48.7 | 38.0 - 59.4 | mg/dL |
| | | | 15.7 | 12.3 - 19.2 | mmol/L |
| K Potassium | ISE direct | | 91.5 | 75.9 - 107 | mmol/L |
| | | | 358 | 297 - 419 | mg/dL |
| Na Sodium | ISE direct | | 202 | 176 - 228 | mmol/L |
| | | | 464 | 404 - 525 | mg/dL |
| TP Total Protein | Biuret (UPRO) | | 96.3 | 73.2 - 119 | mg/dL |
| | | | 963 | 732 - 1194 | mg/L |
| UREA Urea | Urease | | 1580 | 1169 - 1991 | mg/dL |
| | | | 263 | 195 - 331 | mmol/L |
| UA Uric Acid | Uricase | | * | - | mg/dL |
| | | | | - | µmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

ROCHE COBAS C / E / MODULAR



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|------------------------------|------------------|--------------|----------------------------|-----------------|
| ALB Microalbumin | Immunoturbidimetric | | 97.3 9.73 | 70.1 - 125 7.01 - 12.5 | mg/L mg/dL |
| AMY Amylase | Colorimetric | AMYL2 | 252 4.21 | 202 - 302 3.37 - 5.04 | U/L µkat/L |
| BUN Urea N | Urease/GLDH | U-BUN UBUNU | 706 252 | 522 - 890 186 - 317 | mg/dL mmol/L |
| CA Calcium | o-Cresolphthalein Complexone | CA | * | - | mmol/L mg/dL |
| | NM-Bapta | CA2 | 2.84 11.4 | 2.36 - 3.32 9.45 - 13.3 | mmol/L mg/dL |
| CL Chloride | ISE indirect | Cl | 151 535 | 130 - 172 460 - 610 | mmol/L mg/dL |
| CREA Creatinine | Jaffe | CREJ2 | 183 16.2 | 139 - 227 12.3 - 20.1 | mg/dL mmol/L |
| | Enzymatic | CREP2 | 176 15.6 | 134 - 218 11.8 - 19.3 | mg/dL mmol/L |
| GLUC Glucose | Hexokinase | GLUC2 | 273 15.2 | 213 - 333 11.8 - 18.5 | mg/dL mmol/L |
| | | GLUC3 | 263 14.6 | 205 - 321 11.4 - 17.8 | mg/dL mmol/L |
| MG Magnesium | Chlorophosphonazo III | MG | * | - | mmol/L mg/dL |
| | Xylidyl Blue | MG2 | 5.27 12.8 | 3.90 - 6.64 9.48 - 16.1 | mmol/L mg/dL |
| PO4 Phosphorus | Phosphomolybdate UV | PHOS2 | 44.5 14.4 | 34.7 - 54.3 11.2 - 17.5 | mg/dL mmol/L |
| K Potassium | ISE indirect | K | 81.3 318 | 67.5 - 95.1 264 - 372 | mmol/L mg/dL |
| Na Sodium | ISE indirect | NA | 183 421 | 159 - 207 366 - 475 | mmol/L mg/dL |
| TP Total Protein | Turbidimetric | TPUC3 | 52.2 522 | 39.7 - 64.7 397 - 647 | mg/dL mg/L |
| UREA Urea | Urease/GLDH | UREAL URELU | 1513 252 | 1120 - 1906 186 - 317 | mg/dL mmol/L |
| UA Uric Acid | Uricase/Colorimetric | UA2 | 12.2 726 | 9.03 - 15.4 537 - 914 | mg/dL µmol/L |


* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

ROCHE INTEGRA



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|----------------------------------|------------------|-------|-------|-----------------|
| ALB Microalbumin | Immunoturbidimetric | | * | - | mg/L mg/dL |
| AMY Amylase | EPS-G7PNP (Alpha) | | * | - | U/L μkat/L |
| BUN Urea N | Urease | | * | - | mg/dL mmol/L |
| CA Calcium | o-Cresolphthalein Complexone | | * | - | mmol/L mg/dL |
| | NM-Bapta (2nd Gen.) | | * | - | mmol/L mg/dL |
| CL Chloride | ISE indirect | | * | - | mmol/L mg/dL |
| CREA Creatinine | Kinetic Alkaline Picrate (Jaffe) | | * | - | mg/dL mmol/L |
| | Enzymatic/PAP | | * | - | mg/dL mmol/L |
| GLUC Glucose | Hexokinase | | * | - | mg/dL mmol/L |
| MG Magnesium | Chlorophosphonazo III | | * | - | mmol/L mg/dL |
| PO4 Phosphorus | Phosphomolybdate-UV | | * | - | mg/dL mmol/L |
| K Potassium | ISE indirect | | * | - | mmol/L mg/dL |
| Na Sodium | ISE indirect | | * | - | mmol/L mg/dL |
| TP Total Protein | Benzathonium Chloride | | * | - | mg/dL mg/L |
| UREA Urea | Urease/GLDH | | * | - | mg/dL mmol/L |
| UA Uric Acid | Uricase | | * | - | mg/dL μmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

SIEMENS ADVIA



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------------|------------------------------------|------------------|-------|--------|--------|
| ALB Microalbumin | PEG-enhanced Immunoturbidimetric | μALB | * | - | mg/L |
| | PEG-enhanced Immunoturbidimetric | μALB_2 | * | - | mg/dL |
| AMY Amylase | G7 PNP | AMYLAS | * | - | U/L |
| | | | | - | μkat/L |
| BUN Urea N | Uricase/Peroxidase | UA | * | - | mg/dL |
| | | UA_c | | - | mmol/L |
| CA Calcium | Arsenazo III | CA_2 | * | - | mmol/L |
| | | CA_2c | | - | mg/dL |
| | CPC for Advia 1200 | CA | * | - | mmol/L |
| | | CA_c | | - | mg/dL |
| CPC for Advia 1650/1800 /2400 | CA | * | - | mmol/L | |
| | CA_c | | - | mg/dL | |
| CL Chloride | ISE indirect | Cl | * | - | mmol/L |
| CREA Creatinine | Jaffe for Advia 1200 | CREA_2 | * | - | mg/dL |
| | | CRE_2c | | - | mmol/L |
| | Jaffe for Advia 1650/1800 /2400 | CRE_2c | * | - | mg/dL |
| | | | | - | mmol/L |
| | Jaffe for Advia 1650/1800 /2400 | CREA_2 | * | - | mg/dL |
| Enzymatic/PAP | ECRE_2 | * | - | mmol/L | |
| GLUC Glucose | Hexokinase | GLUH_c | * | - | mg/dL |
| | | GLUH_3 | | - | mmol/L |
| | Glucose Oxidase Trinder | GLUO | * | - | mg/dL |
| | | GLUO_c | | - | mmol/L |
| MG Magnesium | Xylidyl Blue | MG | * | - | mmol/L |
| | | MG_c | * | - | mg/dL |
| PO4 Phosphorus | Phosphomolybdate/UV | IP | * | - | mg/dL |
| | | IP_c | * | - | mmol/L |
| K Potassium | ISE indirect | | * | - | mmol/L |
| Na Sodium | ISE indirect | | * | - | mg/dL |
| TP Total Protein | Dye Binding | UPRO_2 | * | - | mg/dL |
| UREA Urea | Urease w/ GLDH | UN | * | - | mg/dL |
| | | UN_c | | - | mmol/L |
| UA Uric Acid | Uricase/Peroxidase | UA | * | - | mg/dL |
| | | UA_c | | - | μmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

SIEMENS ATELLICA



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|----------------------------------|------------------|--------------|----------------------------|-----------------|
| ALB Microalbumin | PEG-enhanced Immunoturbidimetric | μALB_2 | 70.6 7.06 | 50.8 - 90.4 5.08 - 9.04 | mg/L mg/dL |
| AMY Amylase | Ed-G7PNP | AMY_2 | 265 4.42 | 212 - 318 3.53 - 5.30 | U/L μkat/L |
| BUN Urea N | Urease/GLDH | UN_c | 712 254 | 527 - 897 188 - 320 | mg/dL mmol/L |
| CA Calcium | Arsenazo III | CA_2 | 2.74 11.0 | 2.27 - 3.21 9.12 - 12.9 | mmol/L mg/dL |
| | o-Cresolphthalein Complexone | Ca | * | - | mmol/L mg/dL |
| CL Chloride | Indirect IMT | | 170 603 | 146 - 194 518 - 687 | mmol/L mg/dL |
| CORT Cortisol | Chemiluminescence | Cor | * | - | μg/dL nmol/L |
| CREA Creatinine | Alkaline Picrate | Crea_2 | 152 13.4 | 116 - 188 10.2 - 16.7 | mg/dL mmol/L |
| | Enzymatic | Ecre_2 | * | - | mg/dL mmol/L |
| GLUC Glucose | Glucose Oxidase | GluO | * | - | mg/dL mmol/L |
| | Hexokinase | GluH_3 | * | - | mg/dL mmol/L |
| MG Magnesium | Xylidyl Blue | Mg | * | - | mmol/L mg/dL |
| PO4 Phosphorus | Phosphomolybdate / UV | IP | 46.5 15.0 | 36.3 - 56.7 11.7 - 18.3 | mg/dL mmol/L |
| K Potassium | Indirect IMT | | 86.8 339 | 72.0 - 102 282 - 397 | mmol/L mg/dL |
| Na Sodium | Indirect IMT | | 180 414 | 157 - 203 360 - 468 | mmol/L mg/dL |
| TP Total Protein | Pyrogallol Red-Molybdate | UCFP | 65.0 650 | 49.4 - 80.6 494 - 806 | mg/dL mg/L |
| UREA Urea | Urease/GLDH | UN_c | 1525 254 | 1129 - 1922 188 - 320 | mg/dL mmol/L |
| UA Uric Acid | Uricase / Peroxidase | UA | 13.2 785 | 9.77 - 16.6 581 - 989 | mg/dL μmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com



TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30



SIEMENS BNII/PROSPEC

Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|---------------|------------------|-------|-------|---------------|
| ALB Microalbumin | Nephelometric | | * | - | mg/L mg/dL |
| TP Total Protein | Nephelometric | BTP | * | - | mg/dL mg/L |

* not available at time of printing

SIEMENS DCA 2000

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|-------------------------|------------------|-------|-------|-----------------|
| ALB Microalbumin | Immunoturbidimetric | | * | - | mg/L mg/dL |
| CREA Creatinine | 3,5 dinitrobenzoic acid | | * | - | mg/dL mmol/L |

BECKMAN COULTER IMAGE

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|--------------------|------------------|-------|-------|---------------|
| ALB Microalbumin | Rate Nephelometric | MA | * | - | mg/L mg/dL |

Flame Photometer


| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|-----------------------|------------------|------------------|-------|-------|-----------------|
| K Potassium | Flame Photometry | | * | - | mmol/L mg/dL |
| Na Sodium | Flame Photometry | | * | - | mmol/L mg/dL |



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

SIEMENS DIMENSION



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|------------------------------|------------------|--------------|----------------------------|-----------------|
| ALB Microalbumin | PETINIA | MALB | 60.0 6.00 | 43.2 - 76.8 4.32 - 7.68 | mg/L mg/dL |
| AMY Amylase | CNPG3 | AMY | 204 3.41 | 163 - 245 2.73 - 4.08 | U/L µkat/L |
| BUN Urea N | Urease/GLDH | URCA | 779 278 | 576 - 981 206 - 350 | mg/dL mmol/L |
| CA Calcium | o-Cresolphthalein Complexone | CA | 3.00 12.0 | 2.49 - 3.51 9.98 - 14.1 | mmol/L mg/dL |
| CL Chloride | Indirect IMT | RXL | 178 631 | 153 - 203 543 - 720 | mmol/L mg/dL |
| | Indirect IMT | Xpand EXL | 183 649 | 157 - 209 558 - 740 | mmol/L mg/dL |
| CREA Creatinine | Modified Jaffe | CRE2 | 175 15.5 | 133 - 217 11.8 - 19.2 | mg/dL mmol/L |
| | Enzymatic | EZCR | * | - | mg/dL mmol/L |
| GLUC Glucose | Hexokinase | GLUC | * | - | mg/dL mmol/L |
| MG Magnesium | Methylthymol Blue | MG | 5.46 13.3 | 4.04 - 6.88 9.82 - 16.7 | mmol/L mg/dL |
| PO4 Phosphorus | Phosphomolybdate | PHOS | 44.3 14.3 | 34.6 - 54.0 11.2 - 17.5 | mg/dL mmol/L |
| K Potassium | Indirect IMT | RXL | 76.9 301 | 63.8 - 90.0 250 - 352 | mmol/L mg/dL |
| | Indirect IMT | Xpand EXL | 83.2 325 | 69.1 - 97.3 270 - 381 | mmol/L mg/dL |
| Na Sodium | Indirect IMT | RXL | 168 386 | 146 - 190 336 - 436 | mmol/L mg/dL |
| | Indirect IMT | Xpand EXL | 177 407 | 154 - 200 354 - 460 | mmol/L mg/dL |
| TP Total Protein | Pyrogallol Red | UCFP | 66.6 666 | 50.6 - 82.6 506 - 826 | mg/dL mg/L |
| UREA Urea | Urease/GLDH | BUN | 1668 278 | 1234 - 2102 206 - 350 | mg/dL mmol/L |
| UA Uric Acid | Uricase | URCA | 13.3 791 | 9.84 - 16.8 585 - 997 | mg/dL µmol/L |


* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740



Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

 2027-04-30

SIEMENS DIMENSION VISTA



Value sheet version: 1

| Short Name / Analyte | Method | Reagent Cat. No. | Value | Range | Unit |
|----------------------------|------------------------------|------------------|----------------------------|--|------------------------------------|
| ALB Microalbumin | Nephelometric | MALB | 205 20.5 | 148 - 262 14.8 - 26.2 | mg/L mg/dL |
| AMY Amylase | CNPG3 | AMY | 215 3.59 | 172 - 258 2.87 - 4.30 | U/L μkat/L |
| BUN Urea N | Urease/GLDH | | 721 257 | 534 - 908 190 - 324 | mg/dL mmol/L |
| CA Calcium | o-Cresolphthalein Complexone | CA | 2.95 11.8 | 2.45 - 3.45 9.82 - 13.8 | mmol/L mg/dL |
| CL Chloride | IMT Indirect | Cl | 186 660 | 160 - 212 567 - 752 | mmol/L mg/dL |
| CREA Creatinine | Modified Jaffe Enzymatic | CRE2 ECREA | 177 15.7 173 15.3 | 135 - 219 11.9 - 19.5 131 - 215 11.6 - 19.0 | mg/dL mmol/L mg/dL mmol/L |
| GLUC Glucose | Hexokinase | GLU | 275 15.2 | 215 - 336 11.9 - 18.5 | mg/dL mmol/L |
| MG Magnesium | Methylthymol Blue | MG | 4.78 11.6 | 3.54 - 6.02 8.60 - 14.6 | mmol/L mg/dL |
| PO4 Phosphorus | Phosphomolybdate | PHOS | 42.5 13.7 | 33.2 - 51.9 10.7 - 16.7 | mg/dL mmol/L |
| K Potassium | IMT Indirect | K | 89.7 351 | 74.5 - 105 291 - 410 | mmol/L mg/dL |
| Na Sodium | IMT Indirect | Na | 153 352 | 133 - 173 306 - 397 | mmol/L mg/dL |
| TP Total Protein | Pyrogallol Red | UCFP | 67.6 676 | 51.4 - 83.8 514 - 838 | mg/dL mg/L |
| UREA Urea | Urease/GLDH | BUN | 1545 257 | 1143 - 1947 190 - 324 | mg/dL mmol/L |
| UA Uric Acid | Uricase | URCA | 14.4 857 | 10.7 - 18.1 634 - 1079 | mg/dL μmol/L |

* not available at time of printing



DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com