TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/ Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric	1 0242	55.5	40.0 - 71.1	mg/L
Microalbumin			5.55	4.00 - 7.11	mg/dL
AMY	EPS G-7	1 0501	213	170 - 255	U/L
Amylase			3.54	2.84 - 4.25	µkat/L
BUN	Urease UV	1 3101	793	587 - 1000	mg/dL
Urea N			283	209 - 356	mmol/L
CAAS	Arsenazo	1 1130	2.66	2.21 - 3.12	mmol/L
Calcium			10.7	8.87 - 12.5	mg/dL
CA	Phosphonazo	1 1181	2.76	2.29 - 3.23	mmol/L
Calcium			11.1	9.18 - 12.9	mg/dL
CREA	Jaffé, without compensation	1 1711	146	111 - 181	mg/dL
Creatinine			12.9	9.78 - 16.0	mmol/L
CREAP	Enzymatic PAP	1 1759	154	117 - 191	mg/dL
Creatinine			13.6	10.3 - 16.9	mmol/L
D	Refractrometer		1.021	1.010 - 1.035	
Specific Gravity					
GLUHK	HK/G6P-DH	1 2511	269	210 - 328	mg/dL
Glucose			14.9	11.6 - 18.2	mmol/L
hCG				Positive	
Human Chorionic Gonadotropin					
MG	Xylidyl blue	1 4610	5.03	3.72 - 6.33	mmol/L
Magnesium			12.2	9.04 - 15.4	mg/dL
OSMO	Freezing point		800	640 - 960	mOsmo /kg
Osmolality					
pH	pH Meter		6.21	5.50 - 6.90	
PO4	Molybdate UV, substrate start	1 5211	49.0	38.2 - 59.7	mg/dL
Phosphorus			15.8	12.3 - 19.3	mmol/L
TP	Pyrogallol red	1 0210	60.1	45.7 - 74.5	mg/dL
Total Protein			601	457 - 745	mg/L
UREA	Urease UV	1 3101	1699	1257 - 2141	mg/dL
Urea			283	209 - 356	mmol/L
UA TBH	TBHBA	1 3021	13.1	9.68 - 16.5	mg/dL
Uric Acid			778	576 - 980	µmol/L
UA	TOOS	1 3001	13.2	9.78 - 16.7	mg/dL
Uric Acid			786	582 - 991	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

respons®910

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric	1 0242	55.5	40.0 - 71.1	mg/L
Microalbumin			5.55	4.00 - 7.11	mg/dL
AMY	EPS G-7	1 0501	213	170 - 255	U/L
Amylase			3.54	2.84 - 4.25	µkat/L
BUN	Urease UV	1 3101	793	587 - 1000	mg/dL
Urea N			283	209 - 356	mmol/L
CA	Phosphonazo	1 1181	2.76	2.29 - 3.23	mmol/L
Calcium			11.1	9.18 - 12.9	mg/dL
CREA	Jaffé , without compensation	1 1711	146	111 - 181	mg/dL
Creatinine			12.9	9.78 - 16.0	mmol/L
CREAP	Enzymatic PAP	1 1759	154	117 - 191	mg/dL
Creatinine			13.6	10.3 - 16.9	mmol/L
GLUHK	HK/G6P-DH	1 2511	269	210 - 328	mg/dL
Glucose			14.9	11.6 - 18.2	mmol/L
MG	Xylidyl blue	1 4610	5.03	3.72 - 6.33	mmol/L
Magnesium			12.2	9.04 - 15.4	mg/dL
PO4	Molybdate UV	1 5211	49.0	38.2 - 59.7	mg/dL
Phosphorus			15.8	12.3 - 19.3	mmol/L
TP	Pyrogallol red	1 0210	60.1	45.7 - 74.5	mg/dL
Total Protein			601	457 - 745	mg/L
UREA	Urease UV	1 3101	1699	1257 - 2141	mg/dL
Urea			283	209 - 356	mmol/L
UA	TOOS	1 3001	13.2	9.78 - 16.7	mg/dL
Uric Acid			786	582 - 991	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

respons®920

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	1 SD	Unit
ALB	Immunoturbidimetric	1 0242	55.5	40.0 - 71.1	5.18	mg/L
Microalbumin			5.55	4.00 - 7.11	0.518	mg/dL
AMY	EPS G-7	1 0501	213	170 - 255	14.2	U/L
Amylase			3.54	2.84 - 4.25	0.236	µkat/L
BUN	Urease UV	1 3101	793	587 - 1000	68.8	mg/dL
Urea N			283	209 - 356	24.5	mmol/L
CA	Phosphonazo	1 1181	2.76	2.29 - 3.23	0.156	mmol/L
Calcium			11.1	9.18 - 12.9	0.627	mg/dL
CL	Direct ISE	960405	146	125 - 166	6.79	mmol/L
Chloride			516	444 - 588	24.1	mg/dL
CREA	Jaffé , without compensation	1 1711	146	111 - 181	11.7	mg/dL
Creatinine			12.9	9.78 - 16.0	1.03	mmol/L
CREAP	Enzymatic PAP	1 1759	154	117 - 191	12.3	mg/dL
Creatinine			13.6	10.3 - 16.9	1.09	µmol/L
GLUHK	HK/G6P-DH	1 2511	269	210 - 328	19.7	mg/dL
Glucose			14.9	11.6 - 18.2	1.09	mmol/L
MG	Xylidyl blue	1 4610	5.03	3.72 - 6.33	0.436	mmol/L
Magnesium			12.2	9.04 - 15.4	1.06	mg/dL
PO4	Molybdate UV	1 5211	49.0	38.2 - 59.7	3.59	mg/dL
Phosphorus			15.8	12.3 - 19.3	1.16	mmol/L
K	Direct ISE	960403	77.8	64.5 - 91.0	4.41	mmol/L
Potassium			304	252 - 356	17.2	mg/dL
Na	Direct ISE	960402	163	142 - 185	7.08	mmol/L
Sodium			376	327 - 424	16.3	mg/dL
UREA	Urease UV	1 3101	1699	1257 - 2141	147	mg/dL
Urea			283	209 - 356	24.5	mmol/L
UA	TOOS	1 3001	13.2	9.78 - 16.7	1.15	mg/dL
Uric Acid			786	582 - 991	68.1	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

respons®940

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	1 SD	Unit
ALB	Immunoturbidimetric	1 0242	55.5	40.0 - 71.1	5.18	mg/L
Microalbumin			5.55	4.00 - 7.11	0.518	mg/dL
AMY	EPS G-7	1 0501	213	170 - 255	14.2	U/L
Amylase			3.54	2.84 - 4.25	0.236	µkat/L
BUN	Urease UV	1 3101	793	587 - 1000	68.8	mg/dL
Urea N			283	209 - 356	24.5	mmol/L
CA	Phosphonazo	1 1181	2.76	2.29 - 3.23	0.156	mmol/L
Calcium			11.1	9.18 - 12.9	0.627	mg/dL
CL	Direct ISE	960405	146	125 - 166	6.79	mmol/L
Chloride			516	444 - 588	24.1	mg/dL
CREA	Jaffé , without compensation	1 1711	146	111 - 181	11.7	mg/dL
Creatinine			12.9	9.78 - 16.0	1.03	mmol/L
CREAP	Enzymatic PAP	1 1759	154	117 - 191	12.3	mg/dL
Creatinine			13.6	10.3 - 16.9	1.09	mmol/L
GLUHK	HK/G6P-DH	1 2511	269	210 - 328	19.7	mg/dL
Glucose			14.9	11.6 - 18.2	1.09	mmol/L
MG	Xylidyl blue	1 4610	5.03	3.72 - 6.33	0.436	mmol/L
Magnesium			12.2	9.04 - 15.4	1.06	mg/dL
PO4	Molybdate UV	1 5211	49.0	38.2 - 59.7	3.59	mg/dL
Phosphorus			15.8	12.3 - 19.3	1.16	mmol/L
K	Direct ISE	960403	77.8	64.5 - 91.0	4.41	mmol/L
Potassium			304	252 - 356	17.2	mg/dL
Na	Direct ISE	960402	163	142 - 185	7.08	mmol/L
Sodium			376	327 - 424	16.3	mg/dL
TP	Pyrogallol red	1 0210	60.1	45.7 - 74.5	4.81	mg/dL
Total Protein			601	457 - 745	48.1	mg/L
UREA	Urease UV	1 3101	1699	1257 - 2141	147	mg/dL
Urea			283	209 - 356	24.5	mmol/L
UA	TOOS	1 3001	13.2	9.78 - 16.7	1.15	mg/dL
Uric Acid			786	582 - 991	68.1	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

BioMajesty®

CE IVD

Updated sheets: www.diasys-diagnostics.com/service-area/value-sheets/

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	1 SD	Unit
ALB	Immunoturbidimetric	1 0242	55.5	40.0 - 71.1	5.18	mg/L
Microalbumin			5.55	4.00 - 7.11	0.518	mg/dL
AMY	EPS G-7	1 0501	213	170 - 255	14.2	U/L
Amylase			3.54	2.84 - 4.25	0.236	µkat/L
BUN	Urease UV	1 3101	793	587 - 1000	68.8	mg/dL
Urea N			283	209 - 356	24.5	mmol/L
CA	Phosphonazo	1 1181	2.76	2.29 - 3.23	0.156	mmol/L
Calcium			11.1	9.18 - 12.9	0.627	mg/dL
CL	Indirect ISE	980132	168	144 - 191	7.84	mmol/L
Chloride			596	512 - 679	27.8	mg/dL
CREA	Jaffé, without compensation	1 1711	146	111 - 181	11.7	mg/dL
Creatinine			12.9	9.78 - 16.0	1.03	mmol/L
CREAP	Enzymatic PAP	1 1759	154	117 - 191	12.3	mg/dL
Creatinine			13.6	10.3 - 16.9	1.09	mmol/L
GLUHK	HK/G6P-DH	1 2511	269	210 - 328	19.7	mg/dL
Glucose			14.9	11.6 - 18.2	1.09	mmol/L
MG	Xylidyl blue	1 4610	5.03	3.72 - 6.33	0.436	mmol/L
Magnesium			12.2	9.04 - 15.4	1.06	mg/dL
PO4	Molybdate UV	1 5211	49.0	38.2 - 59.7	3.59	mg/dL
Phosphorus			15.8	12.3 - 19.3	1.16	mmol/L
K	Indirect ISE	980131	85.0	70.5 - 99.4	4.81	mmol/L
Potassium			332	276 - 389	18.8	mg/dL
Na	Indirect ISE	980130	184	160 - 208	7.98	mmol/L
Sodium			423	368 - 478	18.3	mg/dL
TP	Pyrogallol red	1 0210	60.1	45.7 - 74.5	4.81	mg/dL
Total Protein			601	457 - 745	48.1	mg/L
UREA	Urease UV	1 3101	1699	1257 - 2141	147	mg/dL
Urea			283	209 - 356	24.5	mmol/L
UA	TOOS	1 3001	13.2	9.78 - 16.7	1.15	mg/dL
Uric Acid			786	582 - 991	68.1	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

ABBOTT AEROSET/ARCHITECT

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Turbidimetric / Immunoturbidimetric	2K98	94.9	68.3 - 121	mg/L
Microalbumin			9.49	6.83 - 12.1	mg/dL
AMY	CNPG3 37° C	7D58	*	-	U/L
Amylase				-	µkat/L
BUN	Urease/GLDH	7D75	*	-	mg/dL
Urea N				-	mmol/L
CA	Arsenazo III	3L79	2.70	2.24 - 3.16	mmol/L
Calcium			10.8	8.99 - 12.7	mg/dL
		4S91	2.55	2.12 - 2.98	mmol/L
			10.2	8.49 - 12.0	mg/dL
CL	ISE Indirect	2P32	147	126 - 168	mmol/L
Chloride			521	448 - 594	mg/dL
CORT	CMIA	8D15	*	-	µg/dL
Cortisol				-	nmol/L
CREA	Alkaline Picrate	3L81	*	-	mg/dL
Creatinine				-	mmol/L
	Enzymatic	8L24	168	128 - 208	mg/dL
			14.9	11.3 - 18.4	mmol/L
GLUC	Hexokinase/G-6-PDH	3L82	280	219 - 342	mg/dL
Glucose			15.6	12.1 - 19.0	mmol/L
MG	Arsenazo III	7D70	*	-	mmol/L
Magnesium				-	mg/dL
	Enzymatic	3P68	*	-	mmol/L
				-	mg/dL
PO4	Phosphomolybdate	7D71	*	-	mg/dL
Phosphorus				-	mmol/L
		4T07	43.8	34.2 - 53.4	mg/dL
			14.1	11.0 - 17.3	mmol/L
K	ISE Indirect	2P32	83.0	68.9 - 97.1	mmol/L
Potassium			325	269 - 380	mg/dL
Na	ISE Indirect	2P32	183	159 - 207	mmol/L
Sodium			421	366 - 475	mg/dL
TP	Benzethonium Chloride	7D79	59.4	45.1 - 73.7	mg/dL
Total Protein			594	451 - 737	mg/L
UREA	Urease/GLDH	7D75	*	-	mg/dL
Urea				-	mmol/L
	Urease/GLDH	4T12	1675	1240 - 2111	mg/dL
			279	206 - 351	mmol/L
UA	Uricase (Uric-U)	4T13	12.5	9.25 - 15.8	mg/dL
Uric Acid			744	550 - 937	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

ABBOTT ALINITY

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Convers	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric	0.1	08P04	93.6	67.4 - 120	mg/L
Microalbumin				9.36	6.74 - 12.0	mg/dL
AMY	Enzymatic CNPG3 37° C	0.0167	07P58	233	186 - 280	U/L
Amylase				3.88	3.11 - 4.66	µkat/L
	Enzymatic CNPG3 37° C, IFCC	0.0167	07P58	*	-	U/L
					-	µkat/L
BUN	Urease/GLDH	0.3563	08P16	709	525 - 893	mg/dL
Urea N				253	187 - 318	mmol/L
CA	Arsenazo III	4.01	07P57	2.67	2.22 - 3.12	mmol/L
Calcium		10		10.7	8.89 - 12.5	mg/dL
CL	ISE Indirect	3.5461	07P53	149	128 - 170	mmol/L
Chloride		10		528	454 - 602	mg/dL
CORT	CMIA	27.59	08P33	18.7	15.0 - 22.4	µg/dL
Cortisol				515	412 - 618	nmol/L
CREA	Alkaline Picrate	0.0884	07P99	166	126 - 206	mg/dL
Creatinine		10		14.6	11.1 - 18.1	mmol/L
	Enzymatic	0.0884	08P01	167	127 - 207	mg/dL
				14.8	11.2 - 18.3	mmol/L
GLUC	Hexokinase/G-6-PDH	0.0555	07P55	283	221 - 345	mg/dL
Glucose		0.01		15.7	12.3 - 19.2	mmol/L
MG	Enzymatic	2.4307	08P19	5.33	3.94 - 6.72	mmol/L
Magnesium				13.0	9.59 - 16.3	mg/dL
PO4	Phosphomolybdate	0.3229	04U03	42.1	32.8 - 51.4	mg/dL
Phosphorus				13.6	10.6 - 16.6	mmol/L
K	ISE Indirect	3.9098	07P53	84.5	70.1 - 98.9	mmol/L
Potassium				330	274 - 387	mg/dL
Na	ISE Indirect	2.299	07P53	185	161 - 209	mmol/L
Sodium				425	370 - 481	mg/dL
TP	Benzethonium Chloride	10	07P59	57.1	43.4 - 70.8	mg/dL
Total Protein				571	434 - 708	mg/L
UREA	Urease/GLDH	0.1665	08P16	1517	1123 - 1911	mg/dL
Urea				253	187 - 318	mmol/L
UA	Uricase	59.48	08P56	13.0	9.62 - 16.4	mg/dL
Uric Acid				773	572 - 974	µmol/L
		59.48	04U09	12.4	9.18 - 15.6	mg/dL
				738	546 - 929	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

CE IVD

BECKMAN COULTER ACCESS

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CORT	Chemiluminescence	Cort	23.9	19.2 - 28.7	µg/dL
Cortisol			661	528 - 793	nmol/L

BECKMAN COULTER DXI

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CORT	Chemiluminescence	Cort	*	-	µg/dL
Cortisol				-	nmol/L

ORTHO-VITROS IMMUNOASSAY SYSTEMS

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CORT	Chemiluminescence		*	-	µg/dL
Cortisol				-	nmol/L

SIEMENS CENTAUR

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CORT	Chemiluminescence		*	-	µg/dL
Cortisol				-	nmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

BECKMAN COULTER AU

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB Microalbumin	Turbidimetric	OSR6167	*	-	mg/L
				-	mg/dL
	Turbidimetric	UALB	106	76.3 - 136	mg/L
			10.6	7.63 - 13.6	mg/dL
AMY Amylase	CNPG3	OSR6x06	182	146 - 218	U/L
			3.04	2.43 - 3.64	µkat/L
	G7PNP/IFCC	OSR6182	*	-	U/L
				-	µkat/L
BUN Urea N	Urease/GLDH	OSR6x34	720	533 - 907	mg/dL
			257	190 - 323	mmol/L
CA Calcium	Arsenazo III	OSR6x117	2.87	2.38 - 3.36	mmol/L
			11.5	9.55 - 13.5	mg/dL
	o-Cresolphthalein Complexone	OSR6x13	*	-	mmol/L
				-	mg/dL
CL Chloride	ISE		164	141 - 187	mmol/L
			582	500 - 663	mg/dL
CREA Creatinine	Enzymatic	OSR61204	*	-	mg/dL
				-	mmol/L
	Modified Jaffe	OSR6x78	143	109 - 177	mg/dL
			12.6	9.6 - 15.7	mmol/L
GLUC Glucose	Hexokinase	OSR6x21	283	221 - 345	mg/dL
			15.7	12.3 - 19.2	mmol/L
MG Magnesium	Xylidyl Blue	OSR6189	5.09	3.77 - 6.41	mmol/L
			12.4	9.16 - 15.6	mg/dL
PO4 Phosphorus	Phosphomolybdate/UV	OSR6x22	43.8	34.2 - 53.4	mg/dL
			14.1	11.0 - 17.3	mmol/L
K Potassium	ISE		83.8	69.6 - 98.0	mmol/L
			328	272 - 383	mg/dL
Na Sodium	ISE		184	160 - 208	mmol/L
			423	368 - 478	mg/dL
TP Total Protein	Pyrogallol Red	OSR6x70	58.8	44.7 - 72.9	mg/dL
			588	447 - 729	mg/L
UREA Urea	Urease/GLDH	OSR6x34	1541	1140 - 1942	mg/dL
			257	190 - 323	mmol/L
UA Uric Acid	Uricase	OSR6x98	15.3	11.3 - 19.3	mg/dL
			910	673 - 1147	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

BECKMAN COULTER DXC/SNYCHRON

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB Microalbumin	Immunoturbidimetric	MA	*	-	mg/L
				-	mg/dL
AMY Amylase	Enzymatic	AMY	*	-	U/L
				-	µkat/L
	Enzymatic	AMY7	*	-	U/L
				-	µkat/L
BUN Urea N	Urease/GLDH	BUN	*	-	mg/dL
				-	mmol/L
CA Calcium	ISE Indirect	CALC	*	-	mmol/L
				-	mg/dL
CL Chloride	ISE Indirect	Cl	*	-	mmol/L
				-	mg/dL
CREA Creatinine	Enzymatic	CR-E	*	-	mg/dL
				-	mmol/L
	Jaffe	CREA / CR-S CREm	*	-	mg/dL
				-	mmol/L
GLUC Glucose	Hexokinase	GLU / GLUH	*	-	mg/dL
				-	mmol/L
	Glucose Oxidase	GLUCm	*	-	mg/dL
				-	mmol/L
MG Magnesium	Calmagite	MG	*	-	mmol/L
				-	mg/dL
PO4 Phosphorus	Phosphomolybdate -UV	PHS	*	-	mg/dL
				-	mmol/L
	Phosphomolybdate -UV	PHOSm	*	-	mg/dL
				-	mmol/L
K Potassium	ISE indirect	K	*	-	mmol/L
				-	mg/dL
Na Sodium	ISE indirect	NA	*	-	mmol/L
				-	mg/dL
TP Total Protein	Timed Endpoint	M-TP	*	-	mg/dL
				-	mg/L
UREA Urea	Urease/GLDH	UREA	*	-	mg/dL
				-	mmol/L
UA Uric Acid	Uricase	URIC	*	-	mg/dL
				-	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

ORTHO-VITROS CHEMISTRY SYSTEMS

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		67.9	48.9 - 86.9	mg/L
Microalbumin			6.79	4.89 - 8.69	mg/dL
	Immunoturbidimetric (DiAgam)		*	-	mg/L
				-	mg/dL
AMY	Amylopectin, Colorimetric		161	129 - 193	U/L
Amylase			2.70	2.16 - 3.22	µkat/L
BUN	Urease		739	547 - 931	mg/dL
Urea N			263	195 - 331	mmol/L
CA	Arsenazo III		2.77	2.30 - 3.24	mmol/L
Calcium			11.1	9.22 - 13.0	mg/dL
CL	ISE direct		143	123 - 163	mmol/L
Chloride			507	436 - 578	mg/dL
CREA	Colorimetric/Enzymatic		170	129 - 211	mg/dL
Creatinine			15.1	11.5 - 18.7	mmol/L
GLUC	Glucose Oxidase		283	221 - 345	mg/dL
Glucose			15.7	12.3 - 19.2	mmol/L
MG	Formazan Dye		5.71	4.23 - 7.19	mmol/L
Magnesium			13.9	10.3 - 17.5	mg/dL
PO4	Phosphomolybdate		48.7	38.0 - 59.4	mg/dL
Phosphorus			15.7	12.3 - 19.2	mmol/L
K	ISE direct		91.5	75.9 - 107	mmol/L
Potassium			358	297 - 419	mg/dL
Na	ISE direct		202	176 - 228	mmol/L
Sodium			464	404 - 525	mg/dL
TP	Biuret (UPRO)		96.3	73.2 - 119	mg/dL
Total Protein			963	732 - 1194	mg/L
UREA	Urease		1580	1169 - 1991	mg/dL
Urea			263	195 - 331	mmol/L
UA	Uricase		*	-	mg/dL
Uric Acid				-	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

ROCHE COBAS C / E / MODULAR

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		97.3	70.1 - 125	mg/L
Microalbumin			9.73	7.01 - 12.5	mg/dL
AMY	Colorimetric	AMYL2	252	202 - 302	U/L
Amylase			4.21	3.37 - 5.04	µkat/L
BUN	Urease/GLDH	U-BUN	706	522 - 890	mg/dL
Urea N		UBUNU	252	186 - 317	mmol/L
CA	o-Cresolphthalein Complexone	CA	*	-	mmol/L
Calcium				-	mg/dL
	NM-Bapta	CA2	2.84	2.36 - 3.32	mmol/L
			11.4	9.45 - 13.3	mg/dL
CL	ISE indirect	Cl	151	130 - 172	mmol/L
Chloride			535	460 - 610	mg/dL
CREA	Jaffe	CREJ2	154	117 - 191	mg/dL
Creatinine			13.6	10.3 - 16.9	mmol/L
	Enzymatic	CREP2	154	117 - 191	mg/dL
			13.6	10.3 - 16.9	mmol/L
GLUC	Hexokinase	GLUC2	273	213 - 333	mg/dL
Glucose			15.2	11.8 - 18.5	mmol/L
		GLUC3	263	205 - 321	mg/dL
			14.6	11.4 - 17.8	mmol/L
MG	Chlorophosphonazo III	MG	*	-	mmol/L
Magnesium				-	mg/dL
	Xylidyl Blue	MG2	5.27	3.90 - 6.64	mmol/L
			12.8	9.48 - 16.1	mg/dL
PO4	Phosphomolybdate UV	PHOS2	44.5	34.7 - 54.3	mg/dL
Phosphorus			14.4	11.2 - 17.5	mmol/L
K	ISE indirect	K	81.3	67.5 - 95.1	mmol/L
Potassium			318	264 - 372	mg/dL
Na	ISE indirect	NA	183	159 - 207	mmol/L
Sodium			421	366 - 475	mg/dL
TP	Turbidimetric	TPUC3	52.2	39.7 - 64.7	mg/dL
Total Protein			522	397 - 647	mg/L
UREA	Urease/GLDH	UREAL	1513	1120 - 1906	mg/dL
Urea		URELU	252	186 - 317	mmol/L
UA	Uricase/Colorimetric	UA2	12.2	9.03 - 15.4	mg/dL
Uric Acid			726	537 - 914	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

ROCHE INTEGRA

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB Microalbumin	Immunoturbidimetric		*	-	mg/L
				-	mg/dL
AMY Amylase	EPS-G7PNP (Alpha)		*	-	U/L
				-	µkat/L
BUN Urea N	Urease		*	-	mg/dL
				-	mmol/L
CA Calcium	o-Cresolphthalein Complexone		*	-	mmol/L
				-	mg/dL
	NM-Bapta (2nd Gen.)		*	-	mmol/L
				-	mg/dL
CL Chloride	ISE indirect		*	-	mmol/L
				-	mg/dL
CREA Creatinine	Kinetic Alkaline Picrate (Jaffe)		*	-	mg/dL
				-	mmol/L
	Enzymatic/PAP		*	-	mg/dL
				-	mmol/L
GLUC Glucose	Hexokinase		*	-	mg/dL
				-	mmol/L
MG Magnesium	Chlorophosphonazo III		*	-	mmol/L
				-	mg/dL
PO4 Phosphorus	Phosphomolybdate-UV		*	-	mg/dL
				-	mmol/L
K Potassium	ISE indirect		*	-	mmol/L
				-	mg/dL
Na Sodium	ISE indirect		*	-	mmol/L
				-	mg/dL
TP Total Protein	Benzathonium Chloride		*	-	mg/dL
				-	mg/L
UREA Urea	Urease/GLDH		*	-	mg/dL
				-	mmol/L
UA Uric Acid	Uricase		*	-	mg/dL
				-	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

SIEMENS ADVIA

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	PEG-enhanced Immunoturbidimetric	µALB	*	-	mg/L
Microalbumin				-	mg/dL
	PEG-enhanced Immunoturbidimetric	µALB_2	*	-	mg/L
				-	mg/dL
AMY	G7 PNP	AMYLAS	*	-	U/L
Amylase				-	µkat/L
BUN	Uricase/Peroxidase	UA	*	-	mg/dL
Urea N		UA_c		-	mmol/L
CA	Arsenazo III	CA_2	*	-	mmol/L
Calcium		CA_2c		-	mg/dL
	CPC	CA	*	-	mmol/L
	for Advia 1200	CA_c		-	mg/dL
	CPC	CA	*	-	mmol/L
	for Advia 1650/1800 /2400	CA_c		-	mg/dL
CL	ISE indirect	Cl	*	-	mmol/L
Chloride				-	mg/dL
CREA	Jaffe	CREA_2	*	-	mg/dL
Creatinine	for Advia 1200	CRE_2c		-	mmol/L
	Jaffe	CRE_2c	*	-	mg/dL
	for Advia 1650/1800 /2400			-	mmol/L
	Jaffe	CREA_2	*	-	mg/dL
	for Advia 1650/1800 /2400			-	mmol/L
GLUC	Hexokinase	GLUH_c	*	-	mg/dL
Glucose		GLUH_3		-	mmol/L
	Glucose Oxidase Trinder	GLUO	*	-	mg/dL
		GLUO_c		-	mmol/L
MG	Xylidyl Blue	MG	*	-	mmol/L
Magnesium				-	mg/dL
	Xylidyl Blue	MG_c	*	-	mmol/L
				-	mg/dL
PO4	Phosphomolybdate/UV	IP	*	-	mg/dL
Phosphorus				-	mmol/L
K	ISE indirect		*	-	mmol/L
Potassium				-	mg/dL
Na	ISE indirect		*	-	mmol/L
Sodium				-	mg/dL
TP	Dye Binding	UPRO_2	*	-	mg/dL
Total Protein				-	mg/L
UREA	Urease w/ GLDH	UN	*	-	mg/dL
Urea		UN_c		-	mmol/L
UA	Uricase/Peroxidase	UA	*	-	mg/dL
Uric Acid		UA_c		-	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

SIEMENS ATELLICA

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	PEG-enhanced Immunoturbidimetric	µALB_2	70.6	50.8 - 90.4	mg/L
Microalbumin			7.06	5.08 - 9.04	mg/dL
AMY	Ed-G7PNP	AMY_2	265	212 - 318	U/L
Amylase			4.42	3.53 - 5.30	µkat/L
BUN	Urease/GLDH	UN_c	712	527 - 897	mg/dL
Urea N			254	188 - 320	mmol/L
CA	Arsenazo III	CA_2	2.74	2.27 - 3.21	mmol/L
Calcium			11.0	9.12 - 12.9	mg/dL
	o-Cresolphthalein Complexone	Ca	3.00	2.49 - 3.51	mmol/L
			12.0	9.98 - 14.1	mg/dL
CL	Indirect IMT		170	146 - 194	mmol/L
Chloride			603	518 - 687	mg/dL
CORT	Chemiluminescence	Cor	*	-	µg/dL
Cortisol				-	nmol/L
CREA	Alkaline Picrate	Crea_2	152	116 - 188	mg/dL
Creatinine			13.4	10.2 - 16.7	mmol/L
	Enzymatic	ECre_2	156	119 - 193	mg/dL
		ECre3	13.8	10.5 - 17.1	mmol/L
GLUC	Glucose Oxidase	GluO	*	-	mg/dL
Glucose				-	mmol/L
	Hexokinase	GluH_3	278	217 - 339	mg/dL
			15.4	12.0 - 18.8	mmol/L
MG	Xylidyl Blue	Mg	5.36	3.97 - 6.75	mmol/L
Magnesium			13.0	9.64 - 16.4	mg/dL
PO4	Phosphomolybdate / UV	IP	46.5	36.3 - 56.7	mg/dL
Phosphorus			15.0	11.7 - 18.3	mmol/L
K	Indirect IMT		86.8	72.0 - 102	mmol/L
Potassium			339	282 - 397	mg/dL
Na	Indirect IMT		180	157 - 203	mmol/L
Sodium			414	360 - 468	mg/dL
TP	Pyrogallol Red-Molybdate	UCFP	65.0	49.4 - 80.6	mg/dL
Total Protein			650	494 - 806	mg/L
UREA	Urease/GLDH	UN_c	1525	1129 - 1922	mg/dL
Urea			254	188 - 320	mmol/L
UA	Uricase / Peroxidase	UA	13.2	9.77 - 16.6	mg/dL
Uric Acid			785	581 - 989	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

CE IVD

SIEMENS BNII/PROSPEC

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Nephelometric		*	-	mg/L
Microalbumin				-	mg/dL
TP	Nephelometric	BTP	*	-	mg/dL
Total Protein				-	mg/L

* not available at time of printing

SIEMENS DCA 2000

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		*	-	mg/L
Microalbumin				-	mg/dL
CREA	3,5 dinitrobenzoic acid		*	-	mg/dL
Creatinine				-	mmol/L

BECKMAN COULTER IMMAGE

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Rate Nephelometric	MA	*	-	mg/L
Microalbumin				-	mg/dL

Flame Photometer

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
K	Flame Photometry		*	-	mmol/L
Potassium				-	mg/dL
Na	Flame Photometry		*	-	mmol/L
Sodium				-	mg/dL

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

SIEMENS DIMENSION

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	PETINIA	MALB	60.0	43.2 - 76.8	mg/L
Microalbumin			6.00	4.32 - 7.68	mg/dL
AMY	CNPG3	AMY	204	163 - 245	U/L
Amylase			3.41	2.73 - 4.08	µkat/L
BUN	Urease/GLDH	URCA	779	576 - 981	mg/dL
Urea N			278	206 - 350	mmol/L
CA	o-Cresolphthalein Complexone	CA	3.00	2.49 - 3.51	mmol/L
Calcium			12.0	9.98 - 14.1	mg/dL
CL	Indirect IMT	RXL	178	153 - 203	mmol/L
Chloride			631	543 - 720	mg/dL
	Indirect IMT	Xpand	183	157 - 209	mmol/L
		EXL	649	558 - 740	mg/dL
CREA	Modified Jaffe	CRE2	175	133 - 217	mg/dL
Creatinine			15.5	11.8 - 19.2	mmol/L
	Enzymatic	EZCR	*	-	mg/dL
				-	mmol/L
GLUC	Hexokinase	GLUC	*	-	mg/dL
Glucose				-	mmol/L
MG	Methylthymol Blue	MG	5.46	4.04 - 6.88	mmol/L
Magnesium			13.3	9.82 - 16.7	mg/dL
PO4	Phosphomolybdate	PHOS	44.3	34.6 - 54.0	mg/dL
Phosphorus			14.3	11.2 - 17.5	mmol/L
K	Indirect IMT	RXL	76.9	63.8 - 90.0	mmol/L
Potassium			301	250 - 352	mg/dL
	Indirect IMT	Xpand	83.2	69.1 - 97.3	mmol/L
		EXL	325	270 - 381	mg/dL
Na	Indirect IMT	RXL	168	146 - 190	mmol/L
Sodium			386	336 - 436	mg/dL
	Indirect IMT	Xpand	177	154 - 200	mmol/L
		EXL	407	354 - 460	mg/dL
TP	Pyrogallol Red	UCFP	66.6	50.6 - 82.6	mg/dL
Total Protein			666	506 - 826	mg/L
UREA	Urease/GLDH	BUN	1668	1234 - 2102	mg/dL
Urea			278	206 - 350	mmol/L
UA	Uricase	URCA	13.3	9.84 - 16.8	mg/dL
Uric Acid			791	585 - 997	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

SIEMENS DIMENSION VISTA

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Nephelometric	MALB	205	148 - 262	mg/L
Microalbumin			20.5	14.8 - 26.2	mg/dL
AMY	CNPG3	AMY	215	172 - 258	U/L
Amylase			3.59	2.87 - 4.30	µkat/L
BUN	Urease/GLDH		721	534 - 908	mg/dL
Urea N			257	190 - 324	mmol/L
CA	o-Cresolphthalein Complexone	CA	2.95	2.45 - 3.45	mmol/L
Calcium			11.8	9.82 - 13.8	mg/dL
CL	IMT Indirect	Cl	186	160 - 212	mmol/L
Chloride			660	567 - 752	mg/dL
CREA	Modified Jaffe	CRE2	177	135 - 219	mg/dL
Creatinine			15.7	11.9 - 19.5	mmol/L
	Enzymatic	ECREA	173	131 - 215	mg/dL
			15.3	11.6 - 19.0	mmol/L
GLUC	Hexokinase	GLU	275	215 - 336	mg/dL
Glucose			15.2	11.9 - 18.5	mmol/L
MG	Methylthymol Blue	MG	4.78	3.54 - 6.02	mmol/L
Magnesium			11.6	8.60 - 14.6	mg/dL
PO4	Phosphomolybdate	PHOS	42.5	33.2 - 51.9	mg/dL
Phosphorus			13.7	10.7 - 16.7	mmol/L
K	IMT Indirect	K	89.7	74.5 - 105	mmol/L
Potassium			351	291 - 410	mg/dL
Na	IMT Indirect	Na	153	133 - 173	mmol/L
Sodium			352	306 - 397	mg/dL
TP	Pyrogallol Red	UCFP	67.6	51.4 - 83.8	mg/dL
Total Protein			676	514 - 838	mg/L
UREA	Urease/GLDH	BUN	1545	1143 - 1947	mg/dL
Urea			257	190 - 324	mmol/L
UA	Uricase	URCA	14.4	10.7 - 18.1	mg/dL
Uric Acid			857	634 - 1079	µmol/L

* not available at time of printing

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

MINDRAY BS SERIES

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		84.6	60.9 - 108	mg/L
Microalbumin			8.46	6.09 - 10.8	mg/dL
	Immunoturbidimetric (Rgt lot # 46025004)		78.8	56.7 - 101	mg/L
			7.88	5.67 - 10.1	mg/dL
AMY	Ed-G7PNP, IFCC		280	224 - 336	U/L
Amylase			4.67	3.73 - 5.60	µkat/L
BUN	Urease/GLDH, UV		> 450	-	mg/dL
Urea N			> 160	-	mmol/L
CA	Arsenazo III		2.79	2.32 - 3.26	mmol/L
Calcium			11.1	9.21 - 13.0	mg/dL
CL	ISE indirect		167	144 - 190	mmol/L
Chloride			592	509 - 675	mg/dL
CORT	CLIA		27.5	22.0 - 33.0	µg/dL
Cortisol			758	606 - 910	nmol/L
CREA	Sarcosine Oxidase		165	125 - 205	mg/dL
Creatinine			14.6	11.1 - 18.1	mmol/L
GLUC	Hexokinase		292	228 - 356	mg/dL
Glucose			16.2	12.6 - 19.8	mmol/L
MG	Xylidyl Blue (Mg II)		> 3.42	-	mmol/L
Magnesium			> 8.32	-	mg/dL
PO4	Phosphomolybdate (P II)		43.4	33.9 - 52.9	mg/dL
Phosphorus			14.0	10.9 - 17.1	mmol/L
K	ISE indirect		84.7	70.3 - 99.1	mmol/L
Potassium			331	275 - 387	mg/dL
Na	ISE indirect		185	161 - 209	mmol/L
Sodium			425	370 - 481	mg/dL
TP	Pyrogallol Red-Molybdate (TPUC)		56.8	43.2 - 70.4	mg/dL
Total Protein			568	432 - 704	mg/L
UREA	Urease/GLDH, UV		> 964	-	mg/dL
Urea			> 160	-	mmol/L
UA	Uricase/Peroxidase		12.8	9.47 - 16.1	mg/dL
Uric Acid			761	563 - 959	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

THERMO FISHER SCIENTIFIC INDIKO SYSTEMS

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
ALB	Immunoturbidimetric		<192	-	mg/L
Microalbumin			< 19.2	-	mg/dL
AMY	IFCC		251	201 - 302	U/L
Amylase			4.19	3.35 - 5.03	µkat/L
BUN	Urease/GLDH		839	621 - 1057	mg/dL
Urea N			300	222 - 378	mmol/L
CA	Arsenazo III		2.73	2.27 - 3.19	mmol/L
Calcium			10.9	9.05 - 12.8	mg/dL
CREA	Jaffé		144	109 - 179	mg/dL
Creatinine			12.8	9.73 - 15.9	mmol/L
GLUC	Hexokinase		282	220 - 343	mg/dL
Glucose			15.6	12.2 - 19.1	mmol/L
PO4	Molybdate, UV		45.8	35.7 - 55.9	mg/dL
Phosphorus			14.8	11.5 - 18.0	mmol/L
TP	Pyrogallol Red-Molybdate		62.8	47.7 - 77.9	mg/dL
Total Protein			628	477 - 779	mg/L
UREA	Urease/GLDH		1796	1329 - 2263	mg/dL
Urea			300	222 - 378	mmol/L
UA	Uricase		14.0	10.4 - 17.6	mg/dL
Uric Acid			833	616 - 1049	µmol/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com

TruLab Urine Level 2

LOT 38740

Cat. No. 5 9180 99 10 061 (6 x 5 mL)
5 9180 99 10 062 (20 x 5 mL)

2027-04-30

THERMO FISHER SCIENTIFIC KONELAB SYSTEMS

CE IVD

Value sheet version: 2

Short Name / Analyte	Method	Reagent Cat. No.	Value	Range	Unit
CREA	Jaffé		136	103 - 169	mg/dL
Creatinine			12.1	9.20 - 15.0	mmol/L
GLUC	Hexokinase		270	211 - 329	mg/dL
Glucose			15.0	11.7 - 18.3	mmol/L
TP	Pyrogallol Red-Molybdate		59.1	44.9 - 73.3	mg/dL
Total Protein			591	449 - 733	mg/L

DiaSys Diagnostic Systems GmbH
Alte Strasse 9 65558 Holzheim Germany
www.diasys-diagnostics.com